

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VOX060424\
 Data File : VX041722.D
 Acq On : 04 Jun 2024 09:58
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Jun 05 04:51:19 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 30 07:36:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	179630	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	258798	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	240788	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	136219	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	104638	47.496	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.000%
35) Dibromofluoromethane	5.385	113	82977	46.374	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.740%
50) Toluene-d8	8.647	98	291820	46.676	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.360%
62) 4-Bromofluorobenzene	11.079	95	114068	49.433	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.860%

Target Compounds				Qvalue		
2) Dichlorodifluoromethane	1.166	85	79517	50.379	ug/l	97
3) Chloromethane	1.300	50	78761	45.735	ug/l	97
4) Vinyl Chloride	1.374	62	84565	47.701	ug/l	99
5) Bromomethane	1.593	94	60404	48.923	ug/l	99
6) Chloroethane	1.660	64	47485	44.315	ug/l	98
7) Trichlorofluoromethane	1.867	101	137182	47.385	ug/l	97
8) Diethyl Ether	2.136	74	57652	51.671	ug/l	91
9) 1,1,2-Trichlorotrifluo...	2.312	101	84010	51.213	ug/l	88
10) Methyl Iodide	2.440	142	101999	49.863	ug/l	91
11) Tert butyl alcohol	3.001	59	109886	281.955	ug/l	# 90
12) 1,1-Dichloroethene	2.306	96	78936	49.297	ug/l	88
13) Acrolein	2.239	56	114585	315.772	ug/l	97
14) Allyl chloride	2.654	41	140982	53.519	ug/l	99
15) Acrylonitrile	3.068	53	270705	283.230	ug/l	99
16) Acetone	2.392	43	278532	294.877	ug/l	98
17) Carbon Disulfide	2.501	76	173618	49.014	ug/l	99
18) Methyl Acetate	2.709	43	121673	57.840	ug/l	100
19) Methyl tert-butyl Ether	3.117	73	295894	55.232	ug/l	99
20) Methylene Chloride	2.782	84	93575	47.462	ug/l	94
21) trans-1,2-Dichloroethene	3.081	96	83140	50.044	ug/l	93
22) Diisopropyl ether	3.763	45	285718	53.073	ug/l	90
23) Vinyl Acetate	3.721	43	1389164	287.228	ug/l	96
24) 1,1-Dichloroethane	3.605	63	153878	50.549	ug/l	96
25) 2-Butanone	4.562	43	401757	293.414	ug/l	98
26) 2,2-Dichloropropane	4.464	77	129426	51.821	ug/l	97
27) cis-1,2-Dichloroethene	4.483	96	103196	51.964	ug/l	92
28) Bromochloromethane	4.897	49	58012	44.246	ug/l	# 81
29) Tetrahydrofuran	5.007	42	248303	287.547	ug/l	96
30) Chloroform	5.092	83	165852	51.892	ug/l	98
31) Cyclohexane	5.458	56	130799	50.287	ug/l	95
32) 1,1,1-Trichloroethane	5.373	97	145665	52.008	ug/l	99
36) 1,1-Dichloropropene	5.684	75	108323	49.997	ug/l	97
37) Ethyl Acetate	4.714	43	150879	55.181	ug/l	99
38) Carbon Tetrachloride	5.665	117	127190	51.671	ug/l	97
39) Methylcyclohexane	7.372	83	140702	52.860	ug/l	94
40) Benzene	6.031	78	346536	51.171	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060424\
 Data File : VX041722.D
 Acq On : 04 Jun 2024 09:58
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Jun 05 04:51:19 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 30 07:36:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.922	41	81982	61.307	ug/l	97
42) 1,2-Dichloroethane	6.080	62	135428	53.569	ug/l	96
43) Isopropyl Acetate	6.342	43	232314	56.275	ug/l	98
44) Trichloroethene	7.123	130	93809	50.933	ug/l	89
45) 1,2-Dichloropropane	7.427	63	86165	53.513	ug/l	100
46) Dibromomethane	7.580	93	67645	54.774	ug/l #	81
47) Bromodichloromethane	7.818	83	126021	55.110	ug/l	99
48) Methyl methacrylate	7.696	41	116316	58.672	ug/l	96
49) 1,4-Dioxane	7.665	88	53688	1239.445	ug/l #	84
51) 4-Methyl-2-Pentanone	8.573	43	788458	294.536	ug/l	99
52) Toluene	8.714	92	226566	52.717	ug/l	99
53) t-1,3-Dichloropropene	8.976	75	131239	58.305	ug/l	100
54) cis-1,3-Dichloropropene	8.366	75	139896	56.182	ug/l	95
55) 1,1,2-Trichloroethane	9.153	97	95592	55.177	ug/l	98
56) Ethyl methacrylate	9.116	69	157997	59.749	ug/l	99
57) 1,3-Dichloropropane	9.305	76	155037	53.902	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.238	63	414248	332.746	ug/l	94
59) 2-Hexanone	9.433	43	626899	307.166	ug/l	98
60) Dibromochloromethane	9.518	129	112051	58.303	ug/l	99
61) 1,2-Dibromoethane	9.610	107	101290	55.395	ug/l	100
64) Tetrachloroethene	9.268	164	92921	51.489	ug/l	91
65) Chlorobenzene	10.079	112	262293	50.757	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.159	131	96192	53.644	ug/l	98
67) Ethyl Benzene	10.195	91	433687	52.661	ug/l	97
68) m/p-Xylenes	10.299	106	351445	107.271	ug/l	92
69) o-Xylene	10.640	106	172900	53.049	ug/l	94
70) Styrene	10.652	104	297788	55.318	ug/l	96
71) Bromoform	10.799	173	90732	58.101	ug/l #	99
73) Isopropylbenzene	10.963	105	443219	52.454	ug/l	96
74) N-amyl acetate	10.841	43	217997	58.241	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.213	83	158677	52.915	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	177925	74.861	ug/l	90
77) Bromobenzene	11.195	156	129064	51.747	ug/l	80
78) n-propylbenzene	11.305	91	498791	53.758	ug/l	96
79) 2-Chlorotoluene	11.366	91	304472	51.055	ug/l	93
80) 1,3,5-Trimethylbenzene	11.451	105	376492	53.366	ug/l	96
81) trans-1,4-Dichloro-2-b...	11.018	75	50911	58.382	ug/l	94
82) 4-Chlorotoluene	11.451	91	357295	52.663	ug/l	92
83) tert-Butylbenzene	11.713	119	398608	52.452	ug/l	92
84) 1,2,4-Trimethylbenzene	11.750	105	383355	53.632	ug/l	93
85) sec-Butylbenzene	11.890	105	475901	54.623	ug/l	96
86) p-Isopropyltoluene	12.006	119	425672	55.708	ug/l	94
87) 1,3-Dichlorobenzene	11.969	146	230023	51.640	ug/l	95
88) 1,4-Dichlorobenzene	12.042	146	234151	50.995	ug/l	96
89) n-Butylbenzene	12.329	91	343361	56.763	ug/l	98
90) Hexachloroethane	12.536	117	70685	54.389	ug/l	64
91) 1,2-Dichlorobenzene	12.335	146	240401	52.985	ug/l	96
92) 1,2-Dibromo-3-Chloropr...	12.945	75	36610	60.433	ug/l	64
93) 1,2,4-Trichlorobenzene	13.585	180	167679	56.075	ug/l	94
94) Hexachlorobutadiene	13.725	225	79970	53.980	ug/l	99
95) Naphthalene	13.774	128	518803	58.794	ug/l	100
96) 1,2,3-Trichlorobenzene	13.963	180	170975	55.557	ug/l	95

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060424\
Data File : VX041722.D
Acq On : 04 Jun 2024 09:58
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1


Instrument :
MSVOA_X
ClientSampleId :


Quant Time: Jun 05 04:51:19 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
Quant Title : SW846 8260
QLast Update : Thu May 30 07:36:24 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jun 05 04:51:19 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
Quant Title : SW846 8260
QLast Update : Thu May 30 07:36:24 2024
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.544 min Scan# 71

Delta R.T. -0.006 min

Lab File: VX041722.D

Acq: 04 Jun 2024 09:58

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion:168 Resp: 179630

Ion Ratio Lower Upper

168 100

99 50.6 56.6 85.0#



#2

Dichlorodifluoromethane

Concen: 50.379 ug/l

RT: 1.166 min Scan# 13

Delta R.T. -0.000 min

Lab File: VX041722.D




Acq: 04 Jun 2024 09:58


Tgt Ion: 85 Resp: 79517

Ion Ratio Lower Upper

85 100

87 31.6 16.7 50.1





#3

Chloromethane

Concen: 45.735 ug/l

RT: 1.300 min Scan# 35

Delta R.T. -0.000 min

Lab File: VX041722.D

Acq: 04 Jun 2024 09:58

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 50 Resp: 78761

Ion Ratio Lower Upper

50 100

52 33.8 25.5 38.3



#4

Vinyl Chloride

Concen: 47.701 ug/l

RT: 1.374 min Scan# 47

Delta R.T. -0.000 min

Lab File: VX041722.D




Acq: 04 Jun 2024 09:58


Tgt Ion: 62 Resp: 84565

Ion Ratio Lower Upper

62 100

64 31.3 25.4 38.0









#5
Bromomethane
Concen: 48.923 ug/l
RT: 1.593 min Scan# 81
Delta R.T. -0.006 min
Lab File: VX041722.D
Acq: 04 Jun 2024 09:58




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 94 Resp: 60404
Ion Ratio Lower Upper
94 100
96 92.3 74.6 111.8



#6
Chloroethane
Concen: 44.315 ug/l
RT: 1.660 min Scan# 94
Delta R.T. -0.006 min
Lab File: VX041722.D
Acq: 04 Jun 2024 09:58

Tgt Ion: 64 Resp: 47485
Ion Ratio Lower Upper
64 100
66 32.7 25.1 37.7









#7
Trichlorofluoromethane
Concen: 47.385 ug/l
RT: 1.867 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX041722.D
Acq: 04 Jun 2024 09:58




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:101 Resp: 137182
Ion Ratio Lower Upper
101 100
103 66.6 51.4 77.2



#8
Diethyl Ether
Concen: 51.671 ug/l
RT: 2.136 min Scan# 172
Delta R.T. -0.000 min
Lab File: VX041722.D
Acq: 04 Jun 2024 09:58

Tgt Ion: 74 Resp: 57652
Ion Ratio Lower Upper
74 100
45 97.4 53.3 159.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 51.213 ug/l
RT: 2.312 min Scan# 20
Delta R.T. -0.006 min
Lab File: VX041722.D
Acq: 04 Jun 2024 09:58




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:101 Resp: 84010
Ion Ratio Lower Upper
101 100
85 42.5 37.0 55.6
151 86.0 58.6 87.8



#10
Methyl Iodide
Concen: 49.863 ug/l
RT: 2.440 min Scan# 222
Delta R.T. -0.006 min
Lab File: VX041722.D
Acq: 04 Jun 2024 09:58

Tgt Ion:142 Resp: 101999
Ion Ratio Lower Upper
142 100
127 43.8 41.6 62.4
141 14.4 11.6 17.4









#11
Tert butyl alcohol
Concen: 281.955 ug/l
RT: 3.001 min Scan# 31
Delta R.T. -0.006 min
Lab File: VX041722.D
Acq: 04 Jun 2024 09:58




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 59 Resp: 109886
Ion Ratio Lower Upper
59 100
57 10.9 6.0 9.0#



#12
1,1-Dichloroethene
Concen: 49.297 ug/l
RT: 2.306 min Scan# 200
Delta R.T. -0.000 min
Lab File: VX041722.D
Acq: 04 Jun 2024 09:58

Tgt Ion: 96 Resp: 78936
Ion Ratio Lower Upper
96 100
61 158.7 144.6 216.8
98 63.4 50.4 75.6





#13

Acrolein

Concen: 315.772 ug/l

RT: 2.239 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX041722.D

Acq: 04 Jun 2024 09:58

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 56 Resp: 114585

Ion Ratio Lower Upper

56 100

55 68.7 57.1 85.7



#14

Allyl chloride

Concen: 53.519 ug/l

RT: 2.654 min Scan# 257

Delta R.T. -0.006 min

Lab File: VX041722.D

Acq: 04 Jun 2024 09:58




Tgt Ion: 41 Resp: 140982


Ion Ratio Lower Upper

41 100

39 69.1 55.8 83.6


76 31.4 24.3 36.5








#15
Acrylonitrile
Concen: 283.230 ug/l
RT: 3.068 min Scan# 31
Delta R.T. -0.006 min
Lab File: VX041722.D
Acq: 04 Jun 2024 09:58

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 53 Resp: 270705
Ion Ratio Lower Upper
53 100
52 82.5 66.0 99.0
51 36.2 29.8 44.6



#16
Acetone
Concen: 294.877 ug/l
RT: 2.392 min Scan# 214
Delta R.T. -0.000 min
Lab File: VX041722.D
Acq: 04 Jun 2024 09:58

Tgt Ion: 43 Resp: 278532
Ion Ratio Lower Upper
43 100
58 28.5 21.9 32.9









#17
Carbon Disulfide
Concen: 49.014 ug/l
RT: 2.501 min Scan# 21
Delta R.T. -0.000 min
Lab File: VX041722.D
Acq: 04 Jun 2024 09:58




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 76 Resp: 173618
Ion Ratio Lower Upper
76 100
78 9.1 7.0 10.6



#18
Methyl Acetate
Concen: 57.840 ug/l
RT: 2.709 min Scan# 266
Delta R.T. -0.000 min
Lab File: VX041722.D
Acq: 04 Jun 2024 09:58

Tgt Ion: 43 Resp: 121673
Ion Ratio Lower Upper
43 100
74 22.3 17.8 26.8









#19
Methyl tert-butyl Ether
Concen: 55.232 ug/l
RT: 3.117 min Scan# 311
Delta R.T. -0.000 min
Lab File: VX041722.D
Acq: 04 Jun 2024 09:58




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 73 Resp: 295894
Ion Ratio Lower Upper
73 100
57 21.0 17.3 25.9



#20
Methylene Chloride
Concen: 47.462 ug/l
RT: 2.782 min Scan# 278
Delta R.T. -0.006 min
Lab File: VX041722.D
Acq: 04 Jun 2024 09:58


Tgt Ion: 84 Resp: 93575
Ion Ratio Lower Upper
84 100
49 121.8 103.6 155.4
51 38.0 33.8 50.6
86 63.7 49.4 74.0








#21
trans-1,2-Dichloroethene
Concen: 50.044 ug/l
RT: 3.081 min Scan# 31
Delta R.T. -0.006 min
Lab File: VX041722.D
Acq: 04 Jun 2024 09:58


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion: 96 Resp: 83140
Ion Ratio Lower Upper
96 100
61 137.4 119.2 178.8
98 62.9 48.7 73.1




#22
Diisopropyl ether
Concen: 53.073 ug/l
RT: 3.763 min Scan# 439
Delta R.T. -0.000 min
Lab File: VX041722.D
Acq: 04 Jun 2024 09:58



Tgt Ion: 45 Resp: 285718
Ion Ratio Lower Upper
45 100
43 75.2 69.7 104.5
87 29.0 20.8 31.2
59 11.6 9.0 13.4





#23

Vinyl Acetate

Concen: 287.228 ug/l

RT: 3.721 min Scan# 41

Delta R.T. -0.000 min

Lab File: VX041722.D

Acq: 04 Jun 2024 09:58

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 43 Resp: 1389164

Ion Ratio Lower Upper

43 100

86 11.1 7.8 11.8



#24

1,1-Dichloroethane

Concen: 50.549 ug/l

RT: 3.605 min Scan# 413

Delta R.T. -0.000 min

Lab File: VX041722.D

Acq: 04 Jun 2024 09:58




Tgt Ion: 63 Resp: 153878


Ion Ratio Lower Upper

63 100

98 8.0 3.3 9.8

100 5.2 2.1 6.5









#25
2-Butanone
Concen: 293.414 ug/l
RT: 4.562 min Scan# 51
Delta R.T. -0.006 min
Lab File: VX041722.D
Acq: 04 Jun 2024 09:58




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 401757
Ion Ratio Lower Upper
43 100
72 25.5 19.4 29.2



#26
2,2-Dichloropropane
Concen: 51.821 ug/l
RT: 4.464 min Scan# 554
Delta R.T. -0.006 min
Lab File: VX041722.D
Acq: 04 Jun 2024 09:58


Tgt Ion: 77 Resp: 129426
Ion Ratio Lower Upper
77 100
97 23.1 10.8 32.3








#27
cis-1,2-Dichloroethene
Concen: 51.964 ug/l
RT: 4.483 min Scan# 51
Delta R.T. -0.000 min
Lab File: VX041722.D
Acq: 04 Jun 2024 09:58

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 96 Resp: 103196
Ion Ratio Lower Upper
96 100
61 141.5 0.0 312.2
98 65.7 0.0 129.0



#28
Bromochloromethane
Concen: 44.246 ug/l
RT: 4.897 min Scan# 625
Delta R.T. -0.000 min
Lab File: VX041722.D
Acq: 04 Jun 2024 09:58

Tgt Ion: 49 Resp: 58012
Ion Ratio Lower Upper
49 100
129 2.4 0.0 3.8
130 88.2 57.5 86.3#









#29
Tetrahydrofuran
Concen: 287.547 ug/l
RT: 5.007 min Scan# 644
Delta R.T. -0.006 min
Lab File: VX041722.D
Acq: 04 Jun 2024 09:58




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 42 Resp: 248303
Ion Ratio Lower Upper
42 100
72 45.7 34.2 51.4
71 42.6 32.2 48.2



#30
Chloroform
Concen: 51.892 ug/l
RT: 5.092 min Scan# 657
Delta R.T. -0.000 min
Lab File: VX041722.D
Acq: 04 Jun 2024 09:58


Tgt Ion: 83 Resp: 165852
Ion Ratio Lower Upper
83 100
85 64.3 50.3 75.5








#31
Cyclohexane
Concen: 50.287 ug/l
RT: 5.458 min Scan# 717
Delta R.T. -0.006 min
Lab File: VX041722.D
Acq: 04 Jun 2024 09:58

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:	56	Resp:	130799
Ion	Ratio	Lower	Upper
56	100		
69	29.6	23.4	35.2
84	88.8	65.9	98.9



#32
1,1,1-Trichloroethane
Concen: 52.008 ug/l
RT: 5.373 min Scan# 703
Delta R.T. -0.006 min
Lab File: VX041722.D
Acq: 04 Jun 2024 09:58

Tgt Ion:	97	Resp:	145665
Ion	Ratio	Lower	Upper
97	100		
99	63.9	51.1	76.7
61	43.4	36.4	54.6





#33

1,2-Dichloroethane-d4

Concen: 47.496 ug/l

RT: 5.952 min Scan# 798

Delta R.T. -0.000 min

Lab File: VX041722.D

Acq: 04 Jun 2024 09:58

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 65 Resp: 104638

Ion Ratio Lower Upper

65 100

67 51.4 0.0 100.2



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. -0.000 min

Lab File: VX041722.D

Acq: 04 Jun 2024 09:58




Tgt Ion: 114 Resp: 258798


Ion Ratio Lower Upper

114 100

63 19.8 0.0 45.0

88 15.9 0.0 36.6









#35
Dibromofluoromethane
Concen: 46.374 ug/l
RT: 5.385 min Scan# 705
Delta R.T. -0.000 min
Lab File: VX041722.D
Acq: 04 Jun 2024 09:58




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:113 Resp: 82977
Ion Ratio Lower Upper
113 100
111 102.5 82.9 124.3
192 24.6 13.3 19.9#



#36
1,1-Dichloropropene
Concen: 49.997 ug/l
RT: 5.684 min Scan# 754
Delta R.T. -0.006 min
Lab File: VX041722.D
Acq: 04 Jun 2024 09:58


Tgt Ion: 75 Resp: 108323
Ion Ratio Lower Upper
75 100
110 37.4 17.0 50.9
77 31.3 24.7 37.1








#37
Ethyl Acetate
Concen: 55.181 ug/l
RT: 4.714 min Scan# 595
Delta R.T. -0.006 min
Lab File: VX041722.D
Acq: 04 Jun 2024 09:58

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 43 Resp: 150879
Ion Ratio Lower Upper
43 100
61 13.6 10.6 16.0
70 10.3 8.3 12.5



#38
Carbon Tetrachloride
Concen: 51.671 ug/l
RT: 5.665 min Scan# 751
Delta R.T. -0.006 min
Lab File: VX041722.D
Acq: 04 Jun 2024 09:58

Tgt Ion: 117 Resp: 127190
Ion Ratio Lower Upper
117 100
119 95.1 78.2 117.2
121 31.7 24.4 36.6









#39
Methylcyclohexane
Concen: 52.860 ug/l
RT: 7.372 min Scan# 1031
Delta R.T. -0.006 min
Lab File: VX041722.D
Acq: 04 Jun 2024 09:58




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 83 Resp: 140702
Ion Ratio Lower Upper
83 100
55 76.0 64.6 97.0
98 51.4 36.7 55.1



#40
Benzene
Concen: 51.171 ug/l
RT: 6.031 min Scan# 811
Delta R.T. -0.006 min
Lab File: VX041722.D
Acq: 04 Jun 2024 09:58

Tgt Ion: 78 Resp: 346536
Ion Ratio Lower Upper
78 100
77 24.0 19.3 28.9









#41
Methacrylonitrile
Concen: 61.307 ug/l
RT: 4.922 min Scan# 619
Delta R.T. -0.000 min
Lab File: VX041722.D
Acq: 04 Jun 2024 09:58




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 41	Resp: 81982
Ion Ratio	Lower Upper
41 100	
39 55.5	45.2 67.8
67 70.4	52.6 78.8
52 29.5	22.7 34.1



#42
1,2-Dichloroethane
Concen: 53.569 ug/l
RT: 6.080 min Scan# 819
Delta R.T. -0.006 min
Lab File: VX041722.D
Acq: 04 Jun 2024 09:58

Tgt Ion: 62	Resp: 135428
Ion Ratio	Lower Upper
62 100	
98 9.3	0.0 16.0





#43

Isopropyl Acetate

Concen: 56.275 ug/l

RT: 6.342 min Scan# 862

Delta R.T. -0.000 min

Lab File: VX041722.D

Acq: 04 Jun 2024 09:58

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 43 Resp: 232314

Ion Ratio Lower Upper

43 100

61 19.4 15.2 22.8

87 13.5 9.8 14.6



#44

Trichloroethene

Concen: 50.933 ug/l

RT: 7.123 min Scan# 990

Delta R.T. -0.000 min

Lab File: VX041722.D




Acq: 04 Jun 2024 09:58


Tgt Ion:130 Resp: 93809

Ion Ratio Lower Upper

130 100

95 90.6 0.0 204.4









#45
1,2-Dichloropropane
Concen: 53.513 ug/l
RT: 7.427 min Scan# 1040
Delta R.T. -0.000 min
Lab File: VX041722.D
Acq: 04 Jun 2024 09:58




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 63 Resp: 86165
Ion Ratio Lower Upper
63 100
65 30.5 24.3 36.5



#46
Dibromomethane
Concen: 54.774 ug/l
RT: 7.580 min Scan# 1065
Delta R.T. -0.000 min
Lab File: VX041722.D
Acq: 04 Jun 2024 09:58


Tgt Ion: 93 Resp: 67645
Ion Ratio Lower Upper
93 100
95 82.9 66.6 100.0
174 121.5 69.4 104.0#








#47
Bromodichloromethane
Concen: 55.110 ug/l
RT: 7.818 min Scan# 1104
Delta R.T. -0.006 min
Lab File: VX041722.D
Acq: 04 Jun 2024 09:58

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 83 Resp: 126021
Ion Ratio Lower Upper
83 100
85 63.2 49.8 74.6
127 9.4 7.1 10.7



#48
Methyl methacrylate
Concen: 58.672 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. -0.000 min
Lab File: VX041722.D
Acq: 04 Jun 2024 09:58

Tgt Ion: 41 Resp: 116316
Ion Ratio Lower Upper
41 100
69 79.5 59.2 88.8
39 60.3 47.7 71.5





#49

1,4-Dioxane

Concen: 1239.445 ug/l

RT: 7.665 min Scan# 1080

Delta R.T. -0.006 min

Lab File: VX041722.D

Acq: 04 Jun 2024 09:58

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 88 Resp: 53688

Ion Ratio Lower Upper

88 100

43 29.5 0.0 0.0#

58 61.1 60.1 90.1



#50

Toluene-d8

Concen: 46.676 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. -0.000 min

Lab File: VX041722.D




Acq: 04 Jun 2024 09:58


Tgt Ion: 98 Resp: 291820

Ion Ratio Lower Upper

98 100

100 67.1 53.1 79.7









#51
4-Methyl-2-Pentanone
Concen: 294.536 ug/l
RT: 8.573 min Scan# 1228
Delta R.T. -0.006 min
Lab File: VX041722.D
Acq: 04 Jun 2024 09:58




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 788458
Ion Ratio Lower Upper
43 100
58 38.7 30.4 45.6



#52
Toluene
Concen: 52.717 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. -0.006 min
Lab File: VX041722.D
Acq: 04 Jun 2024 09:58

Tgt Ion: 92 Resp: 226566
Ion Ratio Lower Upper
92 100
91 169.7 134.2 201.2









#53
 t-1,3-Dichloropropene
 Concen: 58.305 ug/l
 RT: 8.976 min Scan# 1294
 Delta R.T. -0.000 min
 Lab File: VX041722.D
 Acq: 04 Jun 2024 09:58




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion: 75 Resp: 131239
 Ion Ratio Lower Upper
 75 100
 77 31.9 25.7 38.5



#54
 cis-1,3-Dichloropropene
 Concen: 56.182 ug/l
 RT: 8.366 min Scan# 1194
 Delta R.T. -0.000 min
 Lab File: VX041722.D
 Acq: 04 Jun 2024 09:58

Tgt Ion: 75 Resp: 139896
 Ion Ratio Lower Upper
 75 100
 77 32.2 25.3 37.9
 39 53.1 46.7 70.1





#55

1,1,2-Trichloroethane

Concen: 55.177 ug/l

RT: 9.153 min Scan# 1323

Delta R.T. -0.000 min


Lab File: VX041722.D

Acq: 04 Jun 2024 09:58

Instrument :

MSVOA_X

ClientSampleId :



Tgt Ion: 97 Resp: 95592




Ion Ratio Lower Upper

97 100

83 82.2 65.8 98.8

85 54.0 41.6 62.4

99 63.7 47.8 71.8



#56

Ethyl methacrylate

Concen: 59.749 ug/l

RT: 9.116 min Scan# 1317

Delta R.T. -0.000 min

Lab File: VX041722.D

Acq: 04 Jun 2024 09:58




Tgt Ion: 69 Resp: 157997


Ion Ratio Lower Upper

69 100

41 69.8 56.8 85.2

39 40.2 32.0 48.0





#57

1,3-Dichloropropane

Concen: 53.902 ug/l

RT: 9.305 min Scan# 1

Delta R.T. -0.000 min


Lab File: VX041722.D

Acq: 04 Jun 2024 09:58

Instrument :

MSVOA_X

ClientSampleId :






Tgt Ion: 76 Resp: 155037

Ion Ratio Lower Upper

76 100

78 32.4 26.1 39.1



#58

2-Chloroethyl Vinyl ether

Concen: 332.746 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. -0.006 min

Lab File: VX041722.D




Acq: 04 Jun 2024 09:58


Tgt Ion: 63 Resp: 414248

Ion Ratio Lower Upper

63 100

106 31.3 22.5 33.7









#59
2-Hexanone
Concen: 307.166 ug/l
RT: 9.433 min Scan# 1369
Delta R.T. -0.000 min
Lab File: VX041722.D
Acq: 04 Jun 2024 09:58




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 626899
Ion Ratio Lower Upper
43 100
58 53.7 26.2 78.5



#60
Dibromochloromethane
Concen: 58.303 ug/l
RT: 9.518 min Scan# 1383
Delta R.T. -0.000 min
Lab File: VX041722.D
Acq: 04 Jun 2024 09:58


Tgt Ion: 129 Resp: 112051
Ion Ratio Lower Upper
129 100
127 76.3 38.8 116.4








#61
1,2-Dibromoethane
Concen: 55.395 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. -0.000 min
Lab File: VX041722.D
Acq: 04 Jun 2024 09:58

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:107 Resp: 101290
Ion Ratio Lower Upper
107 100
109 94.2 75.2 112.8



#62
4-Bromofluorobenzene
Concen: 49.433 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX041722.D
Acq: 04 Jun 2024 09:58

Tgt Ion: 95 Resp: 114068
Ion Ratio Lower Upper
95 100
174 96.2 0.0 135.6
176 93.1 0.0 131.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX041722.D
Acq: 04 Jun 2024 09:58




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:117 Resp: 240788
Ion Ratio Lower Upper
117 100
82 52.1 46.2 69.4
119 32.3 25.5 38.3



#64
Tetrachloroethene
Concen: 51.489 ug/l
RT: 9.268 min Scan# 1342
Delta R.T. -0.000 min
Lab File: VX041722.D
Acq: 04 Jun 2024 09:58

Tgt Ion:164 Resp: 92921
Ion Ratio Lower Upper
164 100
166 122.9 102.7 154.1
129 88.1 80.4 120.6
131 84.4 76.5 114.7









#65
Chlorobenzene
Concen: 50.757 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VX041722.D
Acq: 04 Jun 2024 09:58




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:112 Resp: 262293
Ion Ratio Lower Upper
112 100
114 32.3 25.7 38.5



#66
1,1,1,2-Tetrachloroethane
Concen: 53.644 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. -0.006 min
Lab File: VX041722.D
Acq: 04 Jun 2024 09:58

Tgt Ion:131 Resp: 96192
Ion Ratio Lower Upper
131 100
133 96.2 47.1 141.4
119 62.1 32.2 96.6









#67
Ethyl Benzene
Concen: 52.661 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. -0.000 min
Lab File: VX041722.D
Acq: 04 Jun 2024 09:58




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 91 Resp: 433687
Ion Ratio Lower Upper
91 100
106 32.5 24.9 37.3



#68
m/p-Xylenes
Concen: 107.271 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. -0.000 min
Lab File: VX041722.D
Acq: 04 Jun 2024 09:58

Tgt Ion: 106 Resp: 351445
Ion Ratio Lower Upper
106 100
91 194.0 165.5 248.3





#69

o-Xylene

Concen: 53.049 ug/l

RT: 10.640 min Scan# 1567

Delta R.T. -0.000 min

Lab File: VX041722.D

Acq: 04 Jun 2024 09:58

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion:106 Resp: 172900

Ion Ratio Lower Upper

106 100

91 209.0 108.9 326.8



#70

Styrene

Concen: 55.318 ug/l

RT: 10.652 min Scan# 1569

Delta R.T. -0.000 min

Lab File: VX041722.D

Acq: 04 Jun 2024 09:58




Tgt Ion:104 Resp: 297788


Ion Ratio Lower Upper

104 100

78 48.4 43.6 65.4

103 54.5 43.6 65.4





#71

Bromoform

Concen: 58.101 ug/l

RT: 10.799 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX041722.D

Acq: 04 Jun 2024 09:58

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:173 Resp: 90732

Ion Ratio Lower Upper

173 100

175 48.6 24.1 72.4

254 0.1 0.0 0.0#



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.024 min Scan# 1794

Delta R.T. -0.000 min

Lab File: VX041722.D

Acq: 04 Jun 2024 09:58




Tgt Ion:152 Resp: 136219


Ion Ratio Lower Upper

152 100

115 77.0 43.3 129.9

150 174.1 0.0 345.0





#73

Isopropylbenzene

Concen: 52.454 ug/l

RT: 10.963 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX041722.D

Acq: 04 Jun 2024 09:58

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion:105 Resp: 443219

Ion Ratio Lower Upper

105 100

120 27.5 12.9 38.6



#74

N-ethyl acetate

Concen: 58.241 ug/l

RT: 10.841 min Scan# 1600

Delta R.T. -0.000 min

Lab File: VX041722.D

Acq: 04 Jun 2024 09:58

Tgt Ion: 43 Resp: 217997




Ion Ratio Lower Upper


43 100

70 42.2 32.9 49.3

55 24.1 19.3 28.9

61 23.9 18.8 28.2





#75

1,1,2,2-Tetrachloroethane

Concen: 52.915 ug/l

RT: 11.213 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX041722.D

Acq: 04 Jun 2024 09:58

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 83 Resp: 158677

Ion Ratio Lower Upper

83 100

131 11.1 5.3 15.8

85 65.2 32.0 96.2



#76

1,2,3-Trichloropropane

Concen: 74.861 ug/l

RT: 11.238 min Scan# 1665

Delta R.T. -0.000 min

Lab File: VX041722.D




Acq: 04 Jun 2024 09:58


Tgt Ion: 75 Resp: 177925

Ion Ratio Lower Upper

75 100

77 32.1 19.1 57.5





#77

Bromobenzene

Concen: 51.747 ug/l

RT: 11.195 min Scan# 1658

Delta R.T. -0.000 min

Lab File: VX041722.D

Acq: 04 Jun 2024 09:58

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:156 Resp: 129064

Ion Ratio Lower Upper

156 100

77 134.8 88.4 265.3

158 96.4 48.0 144.0



#78

n-propylbenzene

Concen: 53.758 ug/l

RT: 11.305 min Scan# 1676

Delta R.T. -0.000 min

Lab File: VX041722.D




Acq: 04 Jun 2024 09:58


Tgt Ion: 91 Resp: 498791

Ion Ratio Lower Upper

91 100

120 24.7 11.5 34.4









#79
2-Chlorotoluene
Concen: 51.055 ug/l
RT: 11.366 min Scan# 1686
Delta R.T. -0.000 min
Lab File: VX041722.D
Acq: 04 Jun 2024 09:58




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 91 Resp: 304472
Ion Ratio Lower Upper
91 100
126 36.3 16.3 48.8



#80
1,3,5-Trimethylbenzene
Concen: 53.366 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX041722.D
Acq: 04 Jun 2024 09:58

Tgt Ion: 105 Resp: 376492
Ion Ratio Lower Upper
105 100
120 50.9 24.1 72.3









#81
trans-1,4-Dichloro-2-butene
Concen: 58.382 ug/l
RT: 11.018 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX041722.D
Acq: 04 Jun 2024 09:58




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 75 Resp: 50911
Ion Ratio Lower Upper
75 100
53 105.1 90.5 135.7
89 46.6 36.9 55.3



#82
4-Chlorotoluene
Concen: 52.663 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX041722.D
Acq: 04 Jun 2024 09:58

Tgt Ion: 91 Resp: 357295
Ion Ratio Lower Upper
91 100
126 32.4 14.2 42.6





#83

tert-Butylbenzene

Concen: 52.452 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. -0.000 min

Lab File: VX041722.D

Acq: 04 Jun 2024 09:58

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:119 Resp: 398608

Ion Ratio Lower Upper

119 100

91 54.8 31.2 93.6

134 24.2 11.7 35.0



#84

1,2,4-Trimethylbenzene

Concen: 53.632 ug/l

RT: 11.750 min Scan# 1749

Delta R.T. -0.000 min

Lab File: VX041722.D




Acq: 04 Jun 2024 09:58


Tgt Ion:105 Resp: 383355

Ion Ratio Lower Upper

105 100

120 46.6 21.2 63.6





#85

sec-Butylbenzene

Concen: 54.623 ug/l

RT: 11.890 min Scan# 1772

Delta R.T. -0.000 min

Lab File: VX041722.D

Acq: 04 Jun 2024 09:58

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion:105 Resp: 475901

Ion Ratio Lower Upper

105 100

134 21.5 9.8 29.4



#86

p-Isopropyltoluene

Concen: 55.708 ug/l

RT: 12.006 min Scan# 1791

Delta R.T. -0.000 min

Lab File: VX041722.D

Acq: 04 Jun 2024 09:58




Tgt Ion:119 Resp: 425672


Ion Ratio Lower Upper

119 100

134 27.1 12.8 38.3

91 22.2 13.1 39.3





#87

1,3-Dichlorobenzene

Concen: 51.640 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. -0.000 min

Lab File: VX041722.D

Acq: 04 Jun 2024 09:58

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:146 Resp: 230023

Ion Ratio Lower Upper

146 100

111 37.6 22.1 66.1

148 64.7 31.8 95.3



#88

1,4-Dichlorobenzene

Concen: 50.995 ug/l

RT: 12.042 min Scan# 1797

Delta R.T. -0.000 min

Lab File: VX041722.D

Acq: 04 Jun 2024 09:58




Tgt Ion:146 Resp: 234151


Ion Ratio Lower Upper

146 100

111 36.7 21.3 63.7


148 64.6 31.8 95.4








#89
n-Butylbenzene
Concen: 56.763 ug/l
RT: 12.329 min Scan# 1844
Delta R.T. -0.000 min
Lab File: VX041722.D
Acq: 04 Jun 2024 09:58

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 91 Resp: 343361
Ion Ratio Lower Upper
91 100
92 54.4 27.4 82.0
134 28.6 12.6 37.8



#90
Hexachloroethane
Concen: 54.389 ug/l
RT: 12.536 min Scan# 1878
Delta R.T. -0.000 min
Lab File: VX041722.D
Acq: 04 Jun 2024 09:58

Tgt Ion: 117 Resp: 70685
Ion Ratio Lower Upper
117 100
201 94.4 32.9 98.6









#91
 1,2-Dichlorobenzene
 Concen: 52.985 ug/l
 RT: 12.335 min Scan# 1845
 Delta R.T. -0.000 min
 Lab File: VX041722.D
 Acq: 04 Jun 2024 09:58




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion:146 Resp: 240401
 Ion Ratio Lower Upper
 146 100
 111 38.3 22.4 67.0
 148 63.6 31.6 95.0



#92
 1,2-Dibromo-3-Chloropropane
 Concen: 60.433 ug/l
 RT: 12.945 min Scan# 1945
 Delta R.T. -0.000 min
 Lab File: VX041722.D
 Acq: 04 Jun 2024 09:58

Tgt Ion: 75 Resp: 36610
 Ion Ratio Lower Upper
 75 100
 155 99.1 36.1 108.5
 157 126.4 44.9 134.7





#93

1,2,4-Trichlorobenzene

Concen: 56.075 ug/l

RT: 13.585 min Scan# 2050

Delta R.T. -0.000 min

Lab File: VX041722.D

Acq: 04 Jun 2024 09:58

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:180 Resp: 167679

Ion Ratio Lower Upper

180 100

182 93.9 48.3 144.8

145 28.0 18.0 54.0



#94

Hexachlorobutadiene

Concen: 53.980 ug/l

RT: 13.725 min Scan# 2073

Delta R.T. -0.000 min

Lab File: VX041722.D

Acq: 04 Jun 2024 09:58




Tgt Ion:225 Resp: 79970


Ion Ratio Lower Upper

225 100

223 62.3 31.1 93.2

227 62.9 32.5 97.5





#95

Naphthalene

Concen: 58.794 ug/l

RT: 13.774 min Scan# 2081

Delta R.T. -0.000 min

Lab File: VX041722.D

Acq: 04 Jun 2024 09:58

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:128 Resp: 518803

Ion Ratio Lower Upper

128 100

127 12.9 10.3 15.5

129 10.8 8.8 13.2



#96

1,2,3-Trichlorobenzene

Concen: 55.557 ug/l

RT: 13.963 min Scan# 2112

Delta R.T. -0.000 min

Lab File: VX041722.D

Acq: 04 Jun 2024 09:58

Tgt Ion:180 Resp: 170975

Ion Ratio Lower Upper

180 100

182 96.2 46.9 140.6

145 29.9 17.8 53.5

