

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060424\
 Data File : VX041723.D
 Acq On : 04 Jun 2024 10:25
 Operator : JC/MD
 Sample : VX0604MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0604MBL01

Quant Time: Jun 05 04:52:16 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 30 07:36:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.543	168	152360	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	235353	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	226362	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	111833	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	87635	46.898	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.800%
35) Dibromofluoromethane	5.385	113	74497	45.782	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.560%
50) Toluene-d8	8.646	98	272523	47.931	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.860%
62) 4-Bromofluorobenzene	11.079	95	102701	48.941	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.880%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060424\
Data File : VX041723.D
Acq On : 04 Jun 2024 10:25
Operator : JC/MD
Sample : VX0604MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0604MBL01

Quant Time: Jun 05 04:52:16 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
Quant Title : SW846 8260
QLast Update : Thu May 30 07:36:24 2024
Response via : Initial Calibration

