

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060424\
 Data File : VX041724.D
 Acq On : 04 Jun 2024 10:48
 Operator : JC/MD
 Sample : VX0604WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0604WBL01

Quant Time: Jun 05 04:52:46 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 30 07:36:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	147456	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	226761	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	222600	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	112316	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	84779	46.878	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.760%
35) Dibromofluoromethane	5.385	113	70450	44.935	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.880%
50) Toluene-d8	8.647	98	264910	48.358	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.720%
62) 4-Bromofluorobenzene	11.079	95	101264	50.084	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.160%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060424\
Data File : VX041724.D
Acq On : 04 Jun 2024 10:48
Operator : JC/MD
Sample : VX0604WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0604WBL01

Quant Time: Jun 05 04:52:46 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
Quant Title : SW846 8260
QLast Update : Thu May 30 07:36:24 2024
Response via : Initial Calibration

