

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060424\
 Data File : VX041731.D
 Acq On : 04 Jun 2024 13:31
 Operator : JC/MD
 Sample : P2660-06DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-12DL

Quant Time: Jun 05 04:55:55 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 30 07:36:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	139221	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	211688	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	199624	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	98895	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	79170	46.366	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	92.740%		
35) Dibromofluoromethane	5.391	113	65365	44.660	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	89.320%		
50) Toluene-d8	8.647	98	245696	48.044	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	96.080%		
62) 4-Bromofluorobenzene	11.079	95	91597	48.529	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	97.060%		
Target Compounds					Qvalue	
16) Acetone	2.392	43	9840	10.440	ug/l	96
25) 2-Butanone	4.605	43	3551	3.346	ug/l	# 82
31) Cyclohexane	5.464	56	5794	2.874	ug/l	90
39) Methylcyclohexane	7.379	83	3925	1.803	ug/l	94
40) Benzene	6.037	78	131552	23.749	ug/l	99
52) Toluene	8.720	92	46512	13.231	ug/l	97
67) Ethyl Benzene	10.195	91	260421	38.143	ug/l	97
68) m/p-Xylenes	10.305	106	127596	46.977	ug/l	95
69) o-Xylene	10.640	106	139977	51.804	ug/l	96
73) Isopropylbenzene	10.963	105	7722	1.259	ug/l	94
78) n-propylbenzene	11.305	91	16420	2.438	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	14167	2.766	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	143118	27.579	ug/l	93
95) Naphthalene	13.774	128	69746	10.887	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060424\
Data File : VX041731.D
Acq On : 04 Jun 2024 13:31
Operator : JC/MD
Sample : P2660-06DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-12DL

Quant Time: Jun 05 04:55:55 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
Quant Title : SW846 8260
QLast Update : Thu May 30 07:36:24 2024
Response via : Initial Calibration

