



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060424\
Data File : VX041721.D
Acq On : 04 Jun 2024 08:54
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
Title : SW846 8260
Last Update : Thu May 30 07:36:24 2024



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.5	8870	PASS
75	95	30	60	51.2	23259	PASS
95	95	100	100	100.0	45400	PASS
96	95	5	9	7.2	3255	PASS
173	174	0.00	2	0.3	122	PASS
174	95	50	100	92.9	42157	PASS
175	174	5	9	8.0	3377	PASS
176	174	95	101	99.1	41795	PASS
177	176	5	9	6.6	2775	PASS