

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX060518\
 Data File : VX002317.D
 Acq On : 05 Jun 2018 20:24
 Operator : JC/MD
 Sample : J3268-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS052MW601-180531

Quant Time: Jun 06 06:05:11 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 05 03:22:35 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	221333	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	328549	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	304310	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	162727	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	158408	50.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.10%	
35) Dibromofluoromethane	5.51	113	117213	45.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.74%	
50) Toluene-d8	8.72	98	477440	48.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.42%	
62) 4-Bromofluorobenzene	11.14	95	160192	41.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.90%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.60	96	172582	61.074	ug/l	85
44) Trichloroethene	7.22	130	1901	0.647	ug/l	86
64) Tetrachloroethene	9.34	164	1261	0.402	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX060518\
Data File : VX002317.D
Acq On : 05 Jun 2018 20:24
Operator : JC/MD
Sample : J3268-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS052MW601-180531

Quant Time: Jun 06 06:05:11 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X060418W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 05 03:22:35 2018
Response via : Initial Calibration

