

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX060518\
 Data File : VX002318.D
 Acq On : 05 Jun 2018 20:51
 Operator : JC/MD
 Sample : J3268-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS052MW749-180531

Quant Time: Jun 06 06:08:33 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 05 03:22:35 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	224160	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	330811	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	301623	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	166341	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	162857	50.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.74%	
35) Dibromofluoromethane	5.51	113	117674	45.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.46%	
50) Toluene-d8	8.72	98	476362	47.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.54%	
62) 4-Bromofluorobenzene	11.14	95	164234	42.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.42%	
Target Compounds						
30) Chloroform	5.22	83	4696	0.951	ug/l	89
38) Carbon Tetrachloride	5.79	117	9889	2.718	ug/l	98
44) Trichloroethene	7.22	130	4337	1.465	ug/l	90
64) Tetrachloroethene	9.34	164	56491	18.160	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX060518\
Data File : VX002318.D
Acq On : 05 Jun 2018 20:51
Operator : JC/MD
Sample : J3268-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS052MW749-180531

Quant Time: Jun 06 06:08:33 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X060418W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 05 03:22:35 2018
Response via : Initial Calibration

