

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX060518\
 Data File : VX002339.D
 Acq On : 06 Jun 2018 07:04
 Operator : JC/MD
 Sample : J3309-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-21

Quant Time: Jun 06 08:18:08 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 05 03:22:35 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	234964	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	344200	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	325552	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	199921	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	162833	48.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.70%	
35) Dibromofluoromethane	5.51	113	120196	44.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.80%	
50) Toluene-d8	8.72	98	494804	47.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.38%	
62) 4-Bromofluorobenzene	11.14	95	178635	44.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.30%	
Target Compounds						
39) Methylcyclohexane	7.47	83	1251	0.541	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX060518\
Data File : VX002339.D
Acq On : 06 Jun 2018 07:04
Operator : JC/MD
Sample : J3309-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-21

Quant Time: Jun 06 08:18:08 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X060418W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 05 03:22:35 2018
Response via : Initial Calibration

