

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX060518\
 Data File : VX002331.D
 Acq On : 06 Jun 2018 03:30
 Operator : JC/MD
 Sample : J3308-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-29

Quant Time: Jun 06 07:11:10 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 05 03:22:35 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	218265	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	309825	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	283358	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	156103	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	149890	47.88	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.76%	
35) Dibromofluoromethane	5.51	113	108513	44.54	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.08%	
50) Toluene-d8	8.72	98	443663	47.51	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.02%	
62) 4-Bromofluorobenzene	11.14	95	150850	41.89	ug/l	0.00
Spiked Amount	50.000		Recovery	=	83.78%	


Target Compounds	Qvalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX060518\
Data File : VX002331.D
Acq On : 06 Jun 2018 03:30
Operator : JC/MD
Sample : J3308-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-29

Quant Time: Jun 06 07:11:10 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X060418W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 05 03:22:35 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.67 min Scan# 834

Delta R.T. -0.00 min

Lab File: VX002331.D

Acq: 06 Jun 2018 03:30

Instrument :

MSVOA_X

ClientSampled :


MW-29

Tot Ion:168 Resp: 218265

Ion Ratio Lower Upper




168 100

99 51.7 39.3 58.9



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



#33

1,2-Dichloroethane-d4

Concen: 47.877 ug/l

RT: 6.07 min Scan# 900

Delta R.T. -0.00 min

Lab File: VX002331.D


Acq: 06 Jun 2018 03:30

Tgt Ion: 65 Resp: 149890

Ion Ratio Lower Upper



65 100


67 51.2 0.0 102.6



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.87 min Scan# 1030

Delta R.T. -0.00 min

Lab File: VX002331.D

Acq: 06 Jun 2018 03:30

Instrument :

MSVOA_X

ClientSampleId :

MW-29


Tot Ion:114 Resp: 309825

Ion Ratio Lower Upper

114 100

63 20.7 0.0 41.4




88 14.5 0.0 30.0



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 44.537 ug/l

RT: 5.51 min Scan# 807

Delta R.T. -0.00 min

Lab File: VX002331.D

Acq: 06 Jun 2018 03:30


Tgt Ion:113 Resp: 108513

Ion Ratio Lower Upper

113 100

111 103.0 81.4 122.0



192 23.5 19.1 28.7




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V









#50
Toluene-d8
Concen: 47.507 ug/l
RT: 8.72 min Scan# 1334
Delta R.T. -0.00 min
Lab File: VX002331.D
Acq: 06 Jun 2018 03:30




Instrument :
MSVOA_X
ClientSampleId :
MW-29


Tot Ion: 98 Resp: 443663
Ion Ratio Lower Upper
98 100
100 63.4 50.9 76.3



#62
4-Bromofluorobenzene
Concen: 41.889 ug/l
RT: 11.14 min Scan# 1731
Delta R.T. -0.00 min
Lab File: VX002331.D
Acq: 06 Jun 2018 03:30

Tgt Ion: 95 Resp: 150850
Ion Ratio Lower Upper
95 100
174 94.5 0.0 187.4
176 91.7 0.0 181.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.12 min Scan# 1563
Delta R.T. -0.00 min
Lab File: VX002331.D
Acq: 06 Jun 2018 03:30

Instrument :
MSVOA_X
ClientSampled :
MW-29

Tot Ion:117 Resp: 283358
Ion Ratio Lower Upper
117 100
82 55.9 45.0 67.4
119 32.2 25.8 38.6




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VX0
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.09 min Scan# 1886
Delta R.T. -0.00 min
Lab File: VX002331.D
Acq: 06 Jun 2018 03:30

Tgt Ion:152 Resp: 156103
Ion Ratio Lower Upper
152 100
115 55.3 40.1 120.2
150 158.0 0.0 347.2



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

