

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060524\
 Data File : VX041738.D
 Acq On : 05 Jun 2024 10:18
 Operator : JC/MD
 Sample : VX0605MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0605MBL01

Quant Time: Jun 06 02:09:10 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 30 07:36:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	144045	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	221750	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	211907	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	105490	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	84384	47.765	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.520%
35) Dibromofluoromethane	5.385	113	69621	45.410	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.820%
50) Toluene-d8	8.647	98	255630	47.718	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.440%
62) 4-Bromofluorobenzene	11.079	95	97161	49.141	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.280%
Target Compounds						
16) Acetone	2.392	43	6094	5.167	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060524\
Data File : VX041738.D
Acq On : 05 Jun 2024 10:18
Operator : JC/MD
Sample : VX0605MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0605MBL01

Quant Time: Jun 06 02:09:10 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
Quant Title : SW846 8260
QLast Update : Thu May 30 07:36:24 2024
Response via : Initial Calibration

