

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060524\
 Data File : VX041744.D
 Acq On : 05 Jun 2024 12:55
 Operator : JC/MD
 Sample : P2717-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANK

Quant Time: Jun 06 02:12:43 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 30 07:36:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	148552	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	232535	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	237023	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	134389	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	86496	47.475	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.940%
35) Dibromofluoromethane	5.391	113	73210	45.536	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.080%
50) Toluene-d8	8.646	98	276696	49.255	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.520%
62) 4-Bromofluorobenzene	11.079	95	112224	54.127	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	108.260%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060524\
Data File : VX041744.D
Acq On : 05 Jun 2024 12:55
Operator : JC/MD
Sample : P2717-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK

Quant Time: Jun 06 02:12:43 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
Quant Title : SW846 8260
QLast Update : Thu May 30 07:36:24 2024
Response via : Initial Calibration

