

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX060524\  
 Data File : VX041746.D  
 Acq On : 05 Jun 2024 13:42  
 Operator : JC/MD  
 Sample : P2598-01  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 12 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 SEEPAGE-CHAMBER

Quant Time: Jun 06 02:14:19 2024  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X052924W.M  
 Quant Title : SW846 8260  
 QLast Update : Thu May 30 07:36:24 2024  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|-------|----------|
| -----                       |        |       |          |          |       |          |
| Internal Standards          |        |       |          |          |       |          |
| 1) Pentafluorobenzene       | 5.556  | 168   | 141184   | 50.000   | ug/l  | # 0.00   |
| 34) 1,4-Difluorobenzene     | 6.763  | 114   | 220774   | 50.000   | ug/l  | 0.00     |
| 63) Chlorobenzene-d5        | 10.055 | 117   | 228042   | 50.000   | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.024 | 152   | 135167   | 50.000   | ug/l  | 0.00     |
| System Monitoring Compounds |        |       |          |          |       |          |
| 33) 1,2-Dichloroethane-d4   | 5.958  | 65    | 83414    | 48.172   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 78 - 117 | Recovery | =     | 96.340%  |
| 35) Dibromofluoromethane    | 5.391  | 113   | 70129    | 45.944   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | =     | 91.880%  |
| 50) Toluene-d8              | 8.647  | 98    | 262397   | 49.198   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 92 - 112 | Recovery | =     | 98.400%  |
| 62) 4-Bromofluorobenzene    | 11.079 | 95    | 110309   | 56.037   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 83 - 123 | Recovery | =     | 112.080% |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX060524\  
Data File : VX041746.D  
Acq On : 05 Jun 2024 13:42  
Operator : JC/MD  
Sample : P2598-01  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 12 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
SEEPAGE-CHAMBER

Quant Time: Jun 06 02:14:19 2024  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X052924W.M  
Quant Title : SW846 8260  
QLast Update : Thu May 30 07:36:24 2024  
Response via : Initial Calibration

