

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060524\
 Data File : VX041748.D
 Acq On : 05 Jun 2024 14:29
 Operator : JC/MD
 Sample : P2598-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 I-VAULT

Quant Time: Jun 06 02:15:12 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 30 07:36:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	139501	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	219699	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	229815	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	137794	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	83326	48.702	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.400%
35) Dibromofluoromethane	5.391	113	70994	46.738	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.480%
50) Toluene-d8	8.647	98	261646	49.297	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.600%
62) 4-Bromofluorobenzene	11.079	95	110229	56.271	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	112.540%
Target Compounds						
16) Acetone	2.386	43	5342	4.422	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060524\
Data File : VX041748.D
Acq On : 05 Jun 2024 14:29
Operator : JC/MD
Sample : P2598-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
I-VAULT

Quant Time: Jun 06 02:15:12 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
Quant Title : SW846 8260
QLast Update : Thu May 30 07:36:24 2024
Response via : Initial Calibration

