

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060524\
 Data File : VX041755.D
 Acq On : 05 Jun 2024 17:12
 Operator : JC/MD
 Sample : P2718-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-4B

Quant Time: Jun 06 02:22:43 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 30 07:36:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	135661	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	210617	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	223781	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	132374	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	80748	48.531	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.060%
35) Dibromofluoromethane	5.391	113	67475	46.337	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.680%
50) Toluene-d8	8.647	98	252310	49.588	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.180%
62) 4-Bromofluorobenzene	11.079	95	107394	57.188	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	114.380%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	5432	4.746	ug/l	98
30) Chloroform	5.099	83	4389	1.818	ug/l	96
67) Ethyl Benzene	10.195	91	1905	0.249	ug/l	95
68) m/p-Xylenes	10.305	106	1618	0.531	ug/l	91
84) 1,2,4-Trimethylbenzene	11.756	105	3066	0.441	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060524\
Data File : VX041755.D
Acq On : 05 Jun 2024 17:12
Operator : JC/MD
Sample : P2718-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-4B

Quant Time: Jun 06 02:22:43 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
Quant Title : SW846 8260
QLast Update : Thu May 30 07:36:24 2024
Response via : Initial Calibration

