

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060524\
 Data File : VX041758.D
 Acq On : 05 Jun 2024 18:22
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Jun 06 02:23:52 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 30 07:36:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	138707	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	217535	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	211854	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	104427	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	81629	47.983	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.960%
35) Dibromofluoromethane	5.385	113	69052	45.912	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.820%
50) Toluene-d8	8.647	98	255521	48.622	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.240%
62) 4-Bromofluorobenzene	11.079	95	97036	50.029	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.060%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	3692	2.254	ug/l	# 89
20) Methylene Chloride	2.788	84	1613	1.060	ug/l	# 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060524\
Data File : VX041758.D
Acq On : 05 Jun 2024 18:22
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Jun 06 02:23:52 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
Quant Title : SW846 8260
QLast Update : Thu May 30 07:36:24 2024
Response via : Initial Calibration

