

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060524\
 Data File : VX041744.D
 Acq On : 05 Jun 2024 12:55
 Operator : JC/MD
 Sample : P2717-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANK

Quant Time: Jun 06 02:12:43 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 30 07:36:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	148552	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	232535	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	237023	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	134389	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	86496	47.475	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.940%
35) Dibromofluoromethane	5.391	113	73210	45.536	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.080%
50) Toluene-d8	8.646	98	276696	49.255	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.520%
62) 4-Bromofluorobenzene	11.079	95	112224	54.127	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	108.260%


Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060524\
Data File : VX041744.D
Acq On : 05 Jun 2024 12:55
Operator : JC/MD
Sample : P2717-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK

Quant Time: Jun 06 02:12:43 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
Quant Title : SW846 8260
QLast Update : Thu May 30 07:36:24 2024
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 71
Delta R.T. 0.006 min
Lab File: VX041744.D
Acq: 05 Jun 2024 12:55




Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK


Tgt Ion:168 Resp: 148552
Ion Ratio Lower Upper
168 100
99 47.7 56.6 85.0#



#33
1,2-Dichloroethane-d4
Concen: 47.475 ug/l
RT: 5.958 min Scan# 799
Delta R.T. 0.006 min
Lab File: VX041744.D
Acq: 05 Jun 2024 12:55

Tgt Ion: 65 Resp: 86496
Ion Ratio Lower Upper
65 100
67 50.9 0.0 100.2









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.763 min Scan# 91
Delta R.T. 0.006 min
Lab File: VX041744.D
Acq: 05 Jun 2024 12:55




Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK


Tgt Ion:114 Resp: 232535
Ion Ratio Lower Upper
114 100
63 19.7 0.0 45.0
88 14.8 0.0 36.6



#35
Dibromofluoromethane
Concen: 45.536 ug/l
RT: 5.391 min Scan# 706
Delta R.T. 0.006 min
Lab File: VX041744.D
Acq: 05 Jun 2024 12:55

Tgt Ion:113 Resp: 73210
Ion Ratio Lower Upper
113 100
111 102.0 82.9 124.3
192 24.0 13.3 19.9#









#50
Toluene-d8
Concen: 49.255 ug/l
RT: 8.646 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX041744.D
Acq: 05 Jun 2024 12:55




Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK


Tgt Ion: 98 Resp: 276696
Ion Ratio Lower Upper
98 100
100 65.5 53.1 79.7



#62
4-Bromofluorobenzene
Concen: 54.127 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX041744.D
Acq: 05 Jun 2024 12:55

Tgt Ion: 95 Resp: 112224
Ion Ratio Lower Upper
95 100
174 96.9 0.0 135.6
176 93.6 0.0 131.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX041744.D
Acq: 05 Jun 2024 12:55

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK

Tgt Ion:117 Resp: 237023
Ion Ratio Lower Upper
117 100
82 52.8 46.2 69.4
119 31.9 25.5 38.3



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. -0.000 min
Lab File: VX041744.D
Acq: 05 Jun 2024 12:55

Tgt Ion:152 Resp: 134389
Ion Ratio Lower Upper
152 100
115 54.8 43.3 129.9
150 156.8 0.0 345.0

