

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\WX060618\
 Data File : VX002364.D
 Acq On : 06 Jun 2018 20:22
 Operator : JC/MD
 Sample : J3312-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 W-04

Quant Time: Jun 07 05:08:20 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 05 03:22:35 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	231488	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	332698	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	301401	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	164671	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	157389	47.36	ug/l	0.00
Spiked Amount			Recovery	=	94.72%	
35) Dibromofluoromethane	5.51	113	117708	44.99	ug/l	0.00
Spiked Amount			Recovery	=	89.98%	
50) Toluene-d8	8.72	98	478630	47.73	ug/l	0.00
Spiked Amount			Recovery	=	95.46%	
62) 4-Bromofluorobenzene	11.14	95	164535	42.55	ug/l	0.00
Spiked Amount			Recovery	=	85.10%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
39) Methylcyclohexane	7.46	83	2183	0.77	ug/l	91
73) Isopropylbenzene	11.02	105	61364	5.36	ug/l	100
78) n-propylbenzene	11.37	91	44594	3.50	ug/l	100
83) tert-Butylbenzene	11.77	119	12704	1.33	ug/l	89
85) sec-Butylbenzene	11.95	105	66163	5.90	ug/l	99
89) n-Butylbenzene	12.39	91	19206	2.24	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX060618\
Data File : VX002364.D
Acq On : 06 Jun 2018 20:22
Operator : JC/MD
Sample : J3312-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
W-04

Quant Time: Jun 07 05:08:20 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X060418W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 05 03:22:35 2018
Response via : Initial Calibration

