

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX060618\
 Data File : VX002344.D
 Acq On : 06 Jun 2018 11:05
 Operator : JC/MD
 Sample : VX0606WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0606WBL01

Quant Time: Jun 07 03:56:27 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 05 03:22:35 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	235857	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	350031	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	320776	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	177615	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	191887	57.41	ug/l	0.00
Spiked Amount	50.000		Recovery	=	114.82%	
35) Dibromofluoromethane	5.51	113	143093	51.98	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.96%	
50) Toluene-d8	8.72	98	580834	55.05	ug/l	0.00
Spiked Amount	50.000		Recovery	=	110.10%	
62) 4-Bromofluorobenzene	11.14	95	200050	49.17	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.34%	


Target Compounds	Qvalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX060618\
Data File : VX002344.D
Acq On : 06 Jun 2018 11:05
Operator : JC/MD
Sample : VX0606WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0606WBL01

Quant Time: Jun 07 03:56:27 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X060418W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 05 03:22:35 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.00 ug/l

RT: 5.67 min Scan# 833

Delta R.T. -0.01 min

Lab File: VX002344.D

Acq: 06 Jun 2018 11:05

Instrument :

MSVOA_X

ClientSampleId :


VX0606WBL01

Tot Ion:168 Resp: 235857

Ion Ratio Lower Upper

168 100

99 53.1 39.3 58.9



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.67

80000



60000

40000

20000

0

Time--> 5.50 5.60 5.70 5.80



#33

1,2-Dichloroethane-d4

Concen: 57.41 ug/l

RT: 6.07 min Scan# 899

Delta R.T. -0.01 min

Lab File: VX002344.D


Acq: 06 Jun 2018 11:05

Tgt Ion: 65 Resp: 191887

Ion Ratio Lower Upper

65 100

67 51.4 0.0 102.6



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0

6.07

80000


60000


40000

20000

0

Time--> 6.00 6.10 6.20





#34

1,4-Difluorobenzene

Concen: 50.00 ug/l

RT: 6.87 min Scan# 1030

Delta R.T. -0.00 min

Lab File: VX002344.D

Acq: 06 Jun 2018 11:05

Instrument :

MSVOA_X

ClientSampleId :

VX0606WBL01


Tot Ion:114 Resp: 350031

Ion Ratio Lower Upper

114 100

63 20.2 0.0 41.4




88 14.7 0.0 30.0



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 51.98 ug/l

RT: 5.51 min Scan# 807

Delta R.T. -0.00 min

Lab File: VX002344.D

Acq: 06 Jun 2018 11:05


Tgt Ion:113 Resp: 143093

Ion Ratio Lower Upper

113 100

111 101.8 81.4 122.0



192 23.3 19.1 28.7




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 55.05 ug/l

RT: 8.72 min Scan# 1334

Delta R.T. 0.00 min

Lab File: VX002344.D

Acq: 06 Jun 2018 11:05

Instrument :

MSVOA_X

ClientSampleId :


VX0606WBL01

Tot Ion: 98 Resp: 580834

Ion Ratio Lower Upper




98 100

100 62.9 50.9 76.3



Abundance Ion 98.00 (97.70 to 98.70): VX0

Ion 100.00 (99.70 to 100.70): VX



#62

4-Bromofluorobenzene

Concen: 49.17 ug/l

RT: 11.14 min Scan# 1731

Delta R.T. -0.00 min

Lab File: VX002344.D

Acq: 06 Jun 2018 11:05


Tgt Ion: 95 Resp: 200050

Ion Ratio Lower Upper

95 100

174 93.8 0.0 187.4



176 90.3 0.0 181.0



Abundance Ion 94.90 (94.60 to 95.60): VX0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 10.12 min Scan# 1563
Delta R.T. -0.00 min
Lab File: VX002344.D
Acq: 06 Jun 2018 11:05

Instrument :
MSVOA_X
ClientSampleId :
VX0606WBL01

Tot Ion:117 Resp: 320776
Ion Ratio Lower Upper
117 100
82 54.1 45.0 67.4
119 31.8 25.8 38.6



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 12.09 min Scan# 1886
Delta R.T. -0.00 min
Lab File: VX002344.D
Acq: 06 Jun 2018 11:05

Tgt Ion:152 Resp: 177615
Ion Ratio Lower Upper
152 100
115 55.0 40.1 120.2
150 157.7 0.0 347.2

