

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX060718\
 Data File : VX002374.D
 Acq On : 07 Jun 2018 13:41
 Operator : JC/MD
 Sample : J3340-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LF5-FB-2018-1

Quant Time: Jun 07 16:16:04 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 05 03:22:35 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	244868	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	354004	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	325366	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	184882	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	187998	53.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.92%	
35) Dibromofluoromethane	5.51	113	142113	51.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.10%	
50) Toluene-d8	8.72	98	562346	52.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.40%	
62) 4-Bromofluorobenzene	11.14	95	199531	48.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.98%	
Target Compounds						
20) Methylene Chloride	2.86	84	499134	187.898	ug/l	99
25) 2-Butanone	4.71	43	68268	28.434	ug/l	98
29) Tetrahydrofuran	5.17	42	2628356	1697.940	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX060718\
Data File : VX002374.D
Acq On : 07 Jun 2018 13:41
Operator : JC/MD
Sample : J3340-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LF5-FB-2018-1

Quant Time: Jun 07 16:16:04 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X060418W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 05 03:22:35 2018
Response via : Initial Calibration

