

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX060718\
 Data File : VX002368.D
 Acq On : 07 Jun 2018 09:56
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 08 03:28:29 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 05 03:22:35 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	89	0.00
2 T	Dichlorodifluoromethane	0.520	0.509	2.1	84	0.00
3 P	Chloromethane	0.684	0.566	17.3	81	0.00
4 C	Vinyl Chloride	0.655	0.629	4.0#	87	0.00
5 T	Bromomethane	0.521	0.324	37.8#	77	0.00
6 T	Chloroethane	0.424	0.416	1.9	91	0.00
7 T	Trichlorofluoromethane	0.998	1.050	-5.2	92	0.00
8 T	Diethyl Ether	0.407	0.385	5.4	94	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.539	0.607	-12.6	96	0.00
10 T	Methyl Iodide	0.758	0.501	33.9#	57	0.00
11 T	Tert butyl alcohol	0.160	0.145	9.4	83	-0.01
12 CM	1,1-Dichloroethene	0.532	0.522	1.9#	91	0.00
13 T	Acrolein	0.077	0.055	28.6#	67	0.00
14 T	Allyl chloride	1.155	1.074	7.0	89	0.00
15 T	Acrylonitrile	0.341	0.311	8.8	87	0.00
16 T	Acetone	0.358	0.355	0.8	95	0.00
17 T	Carbon Disulfide	1.403	1.279	8.8	84	0.00
18 T	Methyl Acetate	0.844	0.754	10.7	89	0.00
19 T	Methyl tert-butyl Ether	2.109	1.991	5.6	92	0.00
20 T	Methylene Chloride	0.710	0.601	15.4	92	0.00
21 T	trans-1,2-Dichloroethene	0.587	0.567	3.4	92	0.00
22 T	Diisopropyl ether	2.155	2.132	1.1	97	0.00
23 T	Vinyl Acetate	1.873	1.854	1.0	96	0.00
24 P	1,1-Dichloroethane	1.142	1.106	3.2	92	0.00
25 T	2-Butanone	0.490	0.436	11.0	84	-0.01
26 T	2,2-Dichloropropane	0.892	0.809	9.3	85	0.00
27 T	cis-1,2-Dichloroethene	0.638	0.575	9.9	87	0.00
28 T	Bromochloromethane	0.553	0.505	8.7	85	0.00
29 T	Tetrahydrofuran	0.316	0.268	15.2	81	0.00
30 C	Chloroform	1.102	1.028	6.7#	89	0.00
31 T	Cyclohexane	0.903	0.894	1.0	86	0.00
32 T	1,1,1-Trichloroethane	0.934	0.897	4.0	86	0.00
33 S	1,2-Dichloroethane-d4	0.778	0.666	14.4	81	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	87	0.00
35 S	Dibromofluoromethane	0.393	0.360	8.4	81	0.00
36 T	1,1-Dichloropropene	0.526	0.525	0.2	87	0.00
37 T	Ethyl Acetate	0.632	0.577	8.7	83	0.00
38 T	Carbon Tetrachloride	0.550	0.550	0.0	85	0.00
39 T	Methylcyclohexane	0.599	0.658	-9.8	90	0.00
40 TM	Benzene	1.612	1.559	3.3	87	0.00
41 T	Methacrylonitrile	0.362	0.328	9.4	83	0.00
42 TM	1,2-Dichloroethane	0.640	0.610	4.7	87	0.00
43 T	Isopropyl Acetate	1.074	0.992	7.6	84	0.00
44 TM	Trichloroethene	0.447	0.449	-0.4	87	0.00
45 C	1,2-Dichloropropane	0.435	0.419	3.7#	87	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060718\
 Data File : VX002368.D
 Acq On : 07 Jun 2018 09:56
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 08 03:28:29 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 05 03:22:35 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.295	0.279	5.4	88	0.00
47 T	Bromodichloromethane	0.533	0.521	2.3	85	0.00
48 T	Methyl methacrylate	0.539	0.512	5.0	84	0.00
49 T	1,4-Dioxane	0.010	0.009	10.0	83	0.00
50 S	Toluene-d8	1.507	1.406	6.7	81	0.00
51 T	4-Methyl-2-Pentanone	0.669	0.633	5.4	84	0.00
52 CM	Toluene	1.025	1.039	-1.4#	89	0.00
53 T	t-1,3-Dichloropropene	0.640	0.644	-0.6	87	0.00
54 T	cis-1,3-Dichloropropene	0.604	0.604	0.0	85	0.00
55 T	1,1,2-Trichloroethane	0.422	0.415	1.7	88	0.00
56 T	Ethyl methacrylate	0.673	0.656	2.5	88	0.00
57 T	1,3-Dichloropropane	0.707	0.685	3.1	88	0.00
58 T	2-Chloroethyl Vinyl ether	0.312	0.303	2.9	86	0.00
59 T	2-Hexanone	0.516	0.493	4.5	83	0.00
60 T	Dibromochloromethane	0.424	0.441	-4.0	87	0.00
61 T	1,2-Dibromoethane	0.440	0.438	0.5	87	0.00
62 S	4-Bromofluorobenzene	0.581	0.546	6.0	81	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	88	0.00
64 T	Tetrachloroethene	0.516	0.526	-1.9	89	0.00
65 PM	Chlorobenzene	1.244	1.240	0.3	90	0.00
66 T	1,1,1,2-Tetrachloroethane	0.450	0.459	-2.0	90	0.00
67 C	Ethyl Benzene	2.082	2.150	-3.3#	90	0.00
68 T	m/p-Xylenes	0.809	0.836	-3.3	90	0.00
69 T	o-Xylene	0.805	0.822	-2.1	90	0.00
70 T	Styrene	1.317	1.370	-4.0	90	0.00
71 P	Bromoform	0.357	0.354	0.8	85	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	93	0.00
73 T	Isopropylbenzene	3.478	3.358	3.5	91	0.00
74 T	N-amyl acetate	1.874	1.540	17.8	84	0.00
75 P	1,1,2,2-Tetrachloroethane	1.094	0.952	13.0	89	0.00
76 T	1,2,3-Trichloropropane	1.097	0.968	11.8	87	0.00
77 T	Bromobenzene	0.976	0.906	7.2	91	0.00
78 T	n-propylbenzene	3.866	3.892	-0.7	91	0.00
79 T	2-Chlorotoluene	2.441	2.274	6.8	91	0.00
80 T	1,3,5-Trimethylbenzene	2.983	2.913	2.3	92	0.00
81 T	trans-1,4-Dichloro-2-butene	0.313	0.272	13.1	81	0.00
82 T	4-Chlorotoluene	2.818	2.726	3.3	91	0.00
83 T	tert-Butylbenzene	2.905	2.838	2.3	92	0.00
84 T	1,2,4-Trimethylbenzene	3.071	2.982	2.9	91	0.00
85 T	sec-Butylbenzene	3.406	3.542	-4.0	93	0.00
86 T	p-Isopropyltoluene	3.102	3.227	-4.0	93	0.00
87 T	1,3-Dichlorobenzene	1.746	1.701	2.6	93	0.00
88 T	1,4-Dichlorobenzene	1.800	1.718	4.6	94	0.00
89 T	n-Butylbenzene	2.602	2.830	-8.8	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060718\
Data File : VX002368.D
Acq On : 07 Jun 2018 09:56
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jun 08 03:28:29 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060418W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 05 03:22:35 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.495	0.481	2.8	89	0.00
91 T	1,2-Dichlorobenzene	1.806	1.706	5.5	91	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.239	0.216	9.6	83	0.00
93 T	1,2,4-Trichlorobenzene	1.267	1.292	-2.0	94	0.00
94 T	Hexachlorobutadiene	0.626	0.694	-10.9	96	0.00
95 T	Naphthalene	3.639	3.468	4.7	88	0.00
96 T	1,2,3-Trichlorobenzene	1.292	1.306	-1.1	94	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6