

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX060818\
 Data File : VX002401.D
 Acq On : 08 Jun 2018 14:57
 Operator : JC/MD
 Sample : J3340-07DL 4X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LF5-FB-2018-1DL

Quant Time: Jun 09 02:22:41 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 05 03:22:35 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	229985	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	330507	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	298518	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	163107	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	155637	47.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.24%	
35) Dibromofluoromethane	5.51	113	120015	46.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.36%	
50) Toluene-d8	8.72	98	470856	47.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.52%	
62) 4-Bromofluorobenzene	11.14	95	160577	41.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.60%	
Target Compounds						
20) Methylene Chloride	2.86	84	117473	44.89	ug/l	97
25) 2-Butanone	4.71	43	17289	7.67	ug/l	99
29) Tetrahydrofuran	5.17	42	616644	424.14	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX060818\
Data File : VX002401.D
Acq On : 08 Jun 2018 14:57
Operator : JC/MD
Sample : J3340-07DL 4X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LF5-FB-2018-1DL

Quant Time: Jun 09 02:22:41 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X060418W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 05 03:22:35 2018
Response via : Initial Calibration

