

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX060818\
 Data File : VX002412.D
 Acq On : 08 Jun 2018 19:50
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 09 03:37:49 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 05 03:22:35 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	76	0.00
2 T	Dichlorodifluoromethane	0.520	0.494	5.0	69	0.00
3 P	Chloromethane	0.684	0.561	18.0	68	0.00
4 C	Vinyl Chloride	0.655	0.646	1.4#	75	0.00
5 T	Bromomethane	0.521	0.376	27.8#	76	0.00
6 T	Chloroethane	0.424	0.433	-2.1	80	0.00
7 T	Trichlorofluoromethane	0.998	1.120	-12.2	83	0.00
8 T	Diethyl Ether	0.407	0.401	1.5	83	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.539	0.616	-14.3	83	0.00
10 T	Methyl Iodide	0.758	0.618	18.5	60	0.00
11 T	Tert butyl alcohol	0.160	0.157	1.9	76	0.00
12 CM	1,1-Dichloroethene	0.532	0.552	-3.8#	81	0.00
13 T	Acrolein	0.077	0.076	1.3	80	0.00
14 T	Allyl chloride	1.155	1.094	5.3	76	0.00
15 T	Acrylonitrile	0.341	0.346	-1.5	82	0.00
16 T	Acetone	0.358	0.325	9.2	74	0.00
17 T	Carbon Disulfide	1.403	1.253	10.7	69	0.00
18 T	Methyl Acetate	0.844	0.846	-0.2	85	0.00
19 T	Methyl tert-butyl Ether	2.109	2.091	0.9	82	0.00
20 T	Methylene Chloride	0.710	0.632	11.0	82	0.00
21 T	trans-1,2-Dichloroethene	0.587	0.591	-0.7	81	0.00
22 T	Diisopropyl ether	2.155	1.930	10.4	74	0.00
23 T	Vinyl Acetate	1.873	1.654	11.7	72	0.00
24 P	1,1-Dichloroethane	1.142	0.988	13.5	69	0.00
25 T	2-Butanone	0.490	0.440	10.2	72	0.00
26 T	2,2-Dichloropropane	0.892	0.709	20.5#	63	0.00
27 T	cis-1,2-Dichloroethene	0.638	0.585	8.3	75	0.00
28 T	Bromochloromethane	0.553	0.515	6.9	73	0.00
29 T	Tetrahydrofuran	0.316	0.289	8.5	74	0.00
30 C	Chloroform	1.102	1.047	5.0#	77	0.00
31 T	Cyclohexane	0.903	0.889	1.6	72	0.00
32 T	1,1,1-Trichloroethane	0.934	0.908	2.8	74	0.00
33 S	1,2-Dichloroethane-d4	0.778	0.676	13.1	70	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	75	0.00
35 S	Dibromofluoromethane	0.393	0.363	7.6	71	0.00
36 T	1,1-Dichloropropene	0.526	0.515	2.1	74	0.00
37 T	Ethyl Acetate	0.632	0.599	5.2	74	0.00
38 T	Carbon Tetrachloride	0.550	0.550	0.0	73	0.00
39 T	Methylcyclohexane	0.599	0.620	-3.5	73	0.00
40 TM	Benzene	1.612	1.550	3.8	75	0.00
41 T	Methacrylonitrile	0.362	0.346	4.4	75	0.00
42 TM	1,2-Dichloroethane	0.640	0.610	4.7	75	0.00
43 T	Isopropyl Acetate	1.074	0.999	7.0	73	0.00
44 TM	Trichloroethene	0.447	0.442	1.1	73	0.00
45 C	1,2-Dichloropropane	0.435	0.416	4.4#	74	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX060818\
 Data File : VX002412.D
 Acq On : 08 Jun 2018 19:50
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 09 03:37:49 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 05 03:22:35 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.295	0.278	5.8	76	0.00
47 T	Bromodichloromethane	0.533	0.520	2.4	73	0.00
48 T	Methyl methacrylate	0.539	0.513	4.8	73	0.00
49 T	1,4-Dioxane	0.010	0.010	0.0	75	0.00
50 S	Toluene-d8	1.507	1.386	8.0	69	0.00
51 T	4-Methyl-2-Pentanone	0.669	0.665	0.6	76	0.00
52 CM	Toluene	1.025	1.026	-0.1#	76	0.00
53 T	t-1,3-Dichloropropene	0.640	0.603	5.8	70	0.00
54 T	cis-1,3-Dichloropropene	0.604	0.567	6.1	69	0.00
55 T	1,1,2-Trichloroethane	0.422	0.419	0.7	77	0.00
56 T	Ethyl methacrylate	0.673	0.658	2.2	76	0.00
57 T	1,3-Dichloropropane	0.707	0.692	2.1	77	0.00
58 T	2-Chloroethyl Vinyl ether	0.312	0.305	2.2	74	0.00
59 T	2-Hexanone	0.516	0.506	1.9	74	0.00
60 T	Dibromochloromethane	0.424	0.426	-0.5	73	0.00
61 T	1,2-Dibromoethane	0.440	0.445	-1.1	76	0.00
62 S	4-Bromofluorobenzene	0.581	0.537	7.6	69	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	77	0.00
64 T	Tetrachloroethene	0.516	0.494	4.3	73	0.00
65 PM	Chlorobenzene	1.244	1.211	2.7	77	0.00
66 T	1,1,1,2-Tetrachloroethane	0.450	0.446	0.9	77	0.00
67 C	Ethyl Benzene	2.082	2.082	0.0	76	0.00
68 T	m/p-Xylenes	0.809	0.817	-1.0	77	0.00
69 T	o-Xylene	0.805	0.799	0.7	77	0.00
70 T	Styrene	1.317	1.330	-1.0	76	0.00
71 P	Bromoform	0.357	0.344	3.6	73	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	79	0.00
73 T	Isopropylbenzene	3.478	3.370	3.1	78	0.00
74 T	N-amyl acetate	1.874	1.565	16.5	73	0.00
75 P	1,1,2,2-Tetrachloroethane	1.094	0.997	8.9	79	0.00
76 T	1,2,3-Trichloropropane	1.097	0.986	10.1	75	0.00
77 T	Bromobenzene	0.976	0.908	7.0	78	0.00
78 T	n-propylbenzene	3.866	3.815	1.3	76	0.00
79 T	2-Chlorotoluene	2.441	2.258	7.5	77	0.00
80 T	1,3,5-Trimethylbenzene	2.983	2.884	3.3	78	0.00
81 T	trans-1,4-Dichloro-2-butene	0.313	0.250	20.1#	64	0.00
82 T	4-Chlorotoluene	2.818	2.683	4.8	77	0.00
83 T	tert-Butylbenzene	2.905	2.832	2.5	78	0.00
84 T	1,2,4-Trimethylbenzene	3.071	2.944	4.1	77	0.00
85 T	sec-Butylbenzene	3.406	3.466	-1.8	78	0.00
86 T	p-Isopropyltoluene	3.102	3.112	-0.3	76	0.00
87 T	1,3-Dichlorobenzene	1.746	1.665	4.6	78	0.00
88 T	1,4-Dichlorobenzene	1.800	1.699	5.6	79	0.00
89 T	n-Butylbenzene	2.602	2.681	-3.0	75	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX060818\
Data File : VX002412.D
Acq On : 08 Jun 2018 19:50
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jun 09 03:37:49 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060418W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 05 03:22:35 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.495	0.459	7.3	72	0.00
91 T	1,2-Dichlorobenzene	1.806	1.710	5.3	78	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.239	0.225	5.9	73	0.00
93 T	1,2,4-Trichlorobenzene	1.267	1.227	3.2	77	0.00
94 T	Hexachlorobutadiene	0.626	0.654	-4.5	77	0.00
95 T	Naphthalene	3.639	3.538	2.8	77	0.00
96 T	1,2,3-Trichlorobenzene	1.292	1.266	2.0	78	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 5