

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX060818\
 Data File : VX002408.D
 Acq On : 08 Jun 2018 18:04
 Operator : JC/MD
 Sample : J3407-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE BLANK-SOILREF-6

Quant Time: Jun 09 03:21:02 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 05 03:22:35 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	219252	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	319715	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	291486	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	158804	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	153629	48.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.86%	
35) Dibromofluoromethane	5.51	113	115632	45.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.98%	
50) Toluene-d8	8.72	98	455742	47.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.58%	
62) 4-Bromofluorobenzene	11.14	95	157198	42.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.60%	


Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX060818\
Data File : VX002408.D
Acq On : 08 Jun 2018 18:04
Operator : JC/MD
Sample : J3407-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGEBLANK-SOILREF-6

Quant Time: Jun 09 03:21:02 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X060418W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 05 03:22:35 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.00 ug/l

RT: 5.67 min Scan# 834

Delta R.T. -0.00 min

Lab File: VX002408.D

Acq: 08 Jun 2018 18:04

Instrument :

MSVOA_X

ClientSampleId :


STORAGEBLANK-SOILREF-6

Tgt Ion:168 Resp: 219252

Ion Ratio Lower Upper




168 100

99 50.4 39.3 58.9



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



#33

1,2-Dichloroethane-d4

Concen: 48.93 ug/l

RT: 6.07 min Scan# 900

Delta R.T. 0.00 min

Lab File: VX002408.D


Acq: 08 Jun 2018 18:04

Tgt Ion: 65 Resp: 153629

Ion Ratio Lower Upper



65 100

67 51.2 0.0 102.6



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0





#34

1,4-Difluorobenzene

Concen: 50.00 ug/l

RT: 6.87 min Scan# 1030

Delta R.T. -0.00 min

Lab File: VX002408.D

Acq: 08 Jun 2018 18:04

Instrument :

MSVOA_X

ClientSampleId :

STORAGEBLANK-SOILREF-6


Tgt Ion:114 Resp: 319715

Ion Ratio Lower Upper

114 100

63 21.3 0.0 41.4




88 15.3 0.0 30.0



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 45.99 ug/l

RT: 5.51 min Scan# 808

Delta R.T. 0.01 min

Lab File: VX002408.D

Acq: 08 Jun 2018 18:04


Tgt Ion:113 Resp: 115632

Ion Ratio Lower Upper

113 100

111 100.7 81.4 122.0



192 23.2 19.1 28.7




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 47.29 ug/l

RT: 8.72 min Scan# 1334

Delta R.T. -0.00 min

Lab File: VX002408.D

Acq: 08 Jun 2018 18:04

Instrument :

MSVOA_X

ClientSampleId :


STORAGEBLANK-SOILREF-6

Tgt Ion: 98 Resp: 455742

Ion Ratio Lower Upper

98 100

100 63.2 50.9 76.3



Abundance Ion 98.00 (97.70 to 98.70): VX0

Ion 100.00 (99.70 to 100.70): VX

300000



200000

100000

0

8.60 8.70 8.80

Time-->



#62

4-Bromofluorobenzene

Concen: 42.30 ug/l

RT: 11.14 min Scan# 1731

Delta R.T. -0.00 min

Lab File: VX002408.D

Acq: 08 Jun 2018 18:04


Tgt Ion: 95 Resp: 157198

Ion Ratio Lower Upper

95 100

174 93.7 0.0 187.4

176 91.8 0.0 181.0



Abundance Ion 94.90 (94.60 to 95.60): VX0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V

150000


100000

50000

0

11.05 11.10 11.15 11.20

Time-->









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 10.12 min Scan# 1563
Delta R.T. -0.00 min
Lab File: VX002408.D
Acq: 08 Jun 2018 18:04

Instrument :
MSVOA_X
ClientSampleId :
STORAGEBLANK-SOILREF-6

Tgt Ion:117 Resp: 291486
Ion Ratio Lower Upper
117 100
82 55.4 45.0 67.4
119 31.8 25.8 38.6




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VX0
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 12.09 min Scan# 1886
Delta R.T. -0.00 min
Lab File: VX002408.D
Acq: 08 Jun 2018 18:04

Tgt Ion:152 Resp: 158804
Ion Ratio Lower Upper
152 100
115 55.7 40.1 120.2
150 156.1 0.0 347.2



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

