

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX060918\
 Data File : VX002425.D
 Acq On : 09 Jun 2018 17:02
 Operator : JC/MD
 Sample : J3396-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 W-08

Quant Time: Jun 11 05:00:00 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 05 03:22:35 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	222311	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	324147	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	297859	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	161973	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	153482	48.15	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.30%	
35) Dibromofluoromethane	5.51	113	116562	45.73	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.46%	
50) Toluene-d8	8.72	98	463142	47.40	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.80%	
62) 4-Bromofluorobenzene	11.14	95	159511	42.34	ug/l	0.00
Spiked Amount	50.000		Recovery	=	84.68%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX060918\
Data File : VX002425.D
Acq On : 09 Jun 2018 17:02
Operator : JC/MD
Sample : J3396-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
W-08

Quant Time: Jun 11 05:00:00 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X060418W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 05 03:22:35 2018
Response via : Initial Calibration

