

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX060918\
 Data File : VX002428.D
 Acq On : 09 Jun 2018 18:22
 Operator : JC/MD
 Sample : J3394-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 W-03

Manual Integrations
 APPROVED

MMDadoda
 6/11/2018 2:32:55 PM

Quant Time: Jun 11 08:15:10 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 05 03:22:35 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	209699	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	309382	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	281245	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	156958	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	148985	49.66	ug/l	0.00
Spiked Amount 50.000			Recovery =	99.32%		
35) Dibromofluoromethane	5.51	113	112652	46.30	ug/l	0.00
Spiked Amount 50.000			Recovery =	92.60%		
50) Toluene-d8	8.72	98	443366	47.54	ug/l	0.00
Spiked Amount 50.000			Recovery =	95.08%		
62) 4-Bromofluorobenzene	11.14	95	152754	42.48	ug/l	0.00
Spiked Amount 50.000			Recovery =	84.96%		

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
31) Cyclohexane	5.58	56	16896	4.46	ug/l	95
39) Methylcyclohexane	7.46	83	31531	8.07	ug/l	98
65) Chlorobenzene	10.14	112	11926	1.70	ug/l	97
73) Isopropylbenzene	11.02	105	297529	27.25	ug/l	99
78) n-propylbenzene	11.37	91	282541	23.28	ug/l	99
83) tert-Butylbenzene	11.77	119	25099	2.75	ug/l	94
85) sec-Butylbenzene	11.95	105	171587	16.05	ug/l	85
89) n-Butylbenzene	12.39	91	47523m	5.82	ug/l	
91) 1,2-Dichlorobenzene	12.40	146	6004	1.06	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX060918\
Data File : VX002428.D
Acq On : 09 Jun 2018 18:22
Operator : JC/MD
Sample : J3394-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
W-03

Manual Integrations
APPROVED

MMDadoda
6/11/2018 2:32:55 PM

Quant Time: Jun 11 08:15:10 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060418W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 05 03:22:35 2018
Response via : Initial Calibration

