


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX060918\
Data File : VX002416.D
Acq On : 09 Jun 2018 10:10
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060418W.M
Title : SW846 8260
Last Update : Tue Jun 05 03:22:35 2018



AutoFind: Scans 1730, 1731, 1732; Background Corrected with Scan 1723

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.2	4552	PASS
75	95	30	60	53.7	11536	PASS
95	95	100	100	100.0	21496	PASS
96	95	5	9	7.0	1495	PASS
173	174	0.00	2	1.5	289	PASS
174	95	50	100	88.9	19111	PASS
175	174	5	9	7.9	1513	PASS
176	174	95	101	100.5	19207	PASS
177	176	5	9	6.6	1270	PASS