

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061124\
 Data File : VX041822.D
 Acq On : 11 Jun 2024 11:13
 Operator : JC/MD
 Sample : VX0611WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0611WBL01

Quant Time: Jun 12 05:47:27 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 30 07:36:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	155262	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	245799	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	246339	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	128665	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	93555	49.130	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.260%
35) Dibromofluoromethane	5.385	113	77496	45.601	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.200%
50) Toluene-d8	8.647	98	292543	49.266	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.540%
62) 4-Bromofluorobenzene	11.079	95	112550	51.355	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.700%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061124\
Data File : VX041822.D
Acq On : 11 Jun 2024 11:13
Operator : JC/MD
Sample : VX0611WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0611WBL01

Quant Time: Jun 12 05:47:27 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
Quant Title : SW846 8260
QLast Update : Thu May 30 07:36:24 2024
Response via : Initial Calibration

