

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061124\
 Data File : VX041825.D
 Acq On : 11 Jun 2024 13:10
 Operator : JC/MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VIBLK

Quant Time: Jun 12 05:49:29 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 30 07:36:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	143116	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	224112	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	217742	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	105813	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	85137	48.504	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.000%
35) Dibromofluoromethane	5.385	113	70844	45.721	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.440%
50) Toluene-d8	8.647	98	261502	48.300	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.600%
62) 4-Bromofluorobenzene	11.079	95	96451	48.268	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.540%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061124\
Data File : VX041825.D
Acq On : 11 Jun 2024 13:10
Operator : JC/MD
Sample : VIBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VIBLK

Quant Time: Jun 12 05:49:29 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
Quant Title : SW846 8260
QLast Update : Thu May 30 07:36:24 2024
Response via : Initial Calibration

