

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061124\
 Data File : VX041826.D
 Acq On : 11 Jun 2024 13:47
 Operator : JC/MD
 Sample : P2756-01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OR-02-06062024

Quant Time: Jun 11 14:20:00 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 30 07:36:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	147064	50.000	ug/l	#-0.01
34) 1,4-Difluorobenzene	6.751	114	230478	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	223441	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	110489	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	86294	47.843	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.680%
35) Dibromofluoromethane	5.379	113	68135	42.758	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	85.520%
50) Toluene-d8	8.647	98	271062	48.683	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.360%
62) 4-Bromofluorobenzene	11.079	95	100392	48.852	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.700%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061124\
Data File : VX041826.D
Acq On : 11 Jun 2024 13:47
Operator : JC/MD
Sample : P2756-01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OR-02-06062024

Quant Time: Jun 11 14:20:00 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
Quant Title : SW846 8260
QLast Update : Thu May 30 07:36:24 2024
Response via : Initial Calibration

