

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061124\
 Data File : VX041824.D
 Acq On : 11 Jun 2024 12:46
 Operator : JC/MD
 Sample : VX0611MBSD01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0611MBSD01

Manual Integrations
 APPROVED

Reviewed By : Semsettin
 Yesilyurt

06/12/2024
 Supervised By : Mahesh
 Dadoda

Quant Time: Jun 12 05:48:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 30 07:36:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	173353	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	258690	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	236106	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	124729	50.000	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	5.952	65	105842	49.782	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.560%
35) Dibromofluoromethane	5.385	113	83154	46.492	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.980%
50) Toluene-d8	8.647	98	285015	45.606	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.220%#
62) 4-Bromofluorobenzene	11.079	95	104901	45.479	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.960%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.166	85	28165	18.490	ug/l	95
3) Chloromethane	1.300	50	31303	18.835	ug/l	99
4) Vinyl Chloride	1.373	62	31496	18.409	ug/l	99
5) Bromomethane	1.599	94	21706	18.217	ug/l	99
6) Chloroethane	1.666	64	19999	19.340	ug/l	100
7) Trichlorofluoromethane	1.867	101	56475	20.214	ug/l	97
8) Diethyl Ether	2.136	74	24860	23.088	ug/l	94
9) 1,1,2-Trichlorotrifluo...	2.318	101	33792	21.346	ug/l	88
10) Methyl Iodide	2.446	142	37214	18.851	ug/l	92
11) Tert butyl alcohol	3.001	59	41193	109.524	ug/l	# 92
12) 1,1-Dichloroethene	2.312	96	33867	21.916	ug/l	85
13) Acrolein	2.239	56	40090	114.480	ug/l	99
14) Allyl chloride	2.660	41	50988	20.057	ug/l	99
15) Acrylonitrile	3.068	53	100844	109.330	ug/l	98
16) Acetone	2.392	43	93693	98.696	ug/l	97
17) Carbon Disulfide	2.501	76	54365	15.903	ug/l	97
18) Methyl Acetate	2.709	43	50025	24.642	ug/l	100
19) Methyl tert-butyl Ether	3.117	73	105644	20.434	ug/l	98
20) Methylene Chloride	2.788	84	35491	18.653	ug/l	95
21) trans-1,2-Dichloroethene	3.087	96	29690	18.518	ug/l	91
22) Diisopropyl ether	3.763	45	109785	21.131	ug/l	95
23) Vinyl Acetate	3.721	43	485025	103.917	ug/l	98
24) 1,1-Dichloroethane	3.605	63	58740	19.995	ug/l	98
25) 2-Butanone	4.568	43	142519	107.854	ug/l	98
26) 2,2-Dichloropropane	4.464	77	41776	17.332	ug/l	98
27) cis-1,2-Dichloroethene	4.489	96	37875	19.762	ug/l	93
28) Bromochloromethane	4.897	49	23802	18.811	ug/l	83
29) Tetrahydrofuran	5.013	42	91968	110.360	ug/l	98
30) Chloroform	5.092	83	62620	20.302	ug/l	94
31) Cyclohexane	5.464	56	44011	17.533	ug/l	91
32) 1,1,1-Trichloroethane	5.379	97	52090	19.272	ug/l	99
36) 1,1-Dichloropropene	5.690	75	37267	17.208	ug/l	96
37) Ethyl Acetate	4.720	43	52839	19.333	ug/l	99
38) Carbon Tetrachloride	5.672	117	42969	17.464	ug/l	98
39) Methylcyclohexane	7.379	83	45667	17.164	ug/l	96
40) Benzene	6.037	78	129142	19.078	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061124\
 Data File : VX041824.D
 Acq On : 11 Jun 2024 12:46
 Operator : JC/MD
 Sample : VX0611MBSD01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0611MBSD01

Manual Integrations
 APPROVED

Reviewed By :Semsettin
 Yesilyurt

06/12/2024
 Supervised By :Mahesh
 Dadoda

Quant Time: Jun 12 05:48:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 30 07:36:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.928	41	28991	21.689	ug/l	95
42) 1,2-Dichloroethane	6.086	62	50672	20.052	ug/l	97
43) Isopropyl Acetate	6.342	43	82978	20.109	ug/l	99
44) Trichloroethene	7.122	130	32745	17.786	ug/l	95
45) 1,2-Dichloropropane	7.427	63	32341	20.094	ug/l	99
46) Dibromomethane	7.580	93	24745	20.045	ug/l #	82
47) Bromodichloromethane	7.817	83	44107	19.296	ug/l	98
48) Methyl methacrylate	7.696	41	40989	20.684	ug/l	95
49) 1,4-Dioxane	7.665	88	18334	423.437	ug/l #	93
51) 4-Methyl-2-Pentanone	8.573	43	291666	109.000	ug/l	99
52) Toluene	8.714	92	81476	18.966	ug/l	99
53) t-1,3-Dichloropropene	8.976	75	41602	18.490	ug/l	99
54) cis-1,3-Dichloropropene	8.366	75	46828	18.814	ug/l	95
55) 1,1,2-Trichloroethane	9.153	97	35028	20.227	ug/l	98
56) Ethyl methacrylate	9.116	69	53562	20.264	ug/l	98
57) 1,3-Dichloropropane	9.305	76	58331	20.289	ug/l	98
58) 2-Chloroethyl Vinyl ether	8.238	63	127598	102.536	ug/l	96
59) 2-Hexanone	9.433	43	219981	107.831	ug/l	97
60) Dibromochloromethane	9.518	129	36186	18.836	ug/l	99
61) 1,2-Dibromoethane	9.610	107	36035	19.716	ug/l	100
64) Tetrachloroethene	9.275	164	31698	17.913	ug/l	90
65) Chlorobenzene	10.079	112	93496	18.452	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.159	131	33174	18.867	ug/l	97
67) Ethyl Benzene	10.195	91	149120	18.466	ug/l	98
68) m/p-Xylenes	10.299	106	120591	37.538	ug/l	93
69) o-Xylene	10.640	106	58580	18.330	ug/l	97
70) Styrene	10.652	104	99626	18.874	ug/l	96
71) Bromoform	10.799	173	28862	18.848	ug/l #	100
73) Isopropylbenzene	10.963	105	152319	19.687	ug/l	99
74) N-amyl acetate	10.841	43	71506	20.864	ug/l	98
75) 1,1,2,2-Tetrachloroethane	11.213	83	57942	21.102	ug/l	100
76) 1,2,3-Trichloropropane	11.238	75	46561m	21.395	ug/l	
77) Bromobenzene	11.195	156	44720	19.582	ug/l	81
78) n-propylbenzene	11.305	91	163447	19.239	ug/l	97
79) 2-Chlorotoluene	11.366	91	105627	19.344	ug/l	94
80) 1,3,5-Trimethylbenzene	11.451	105	126418	19.570	ug/l	97
81) trans-1,4-Dichloro-2-b...	11.018	75	14385	18.015	ug/l	97
82) 4-Chlorotoluene	11.451	91	119149	19.180	ug/l	95
83) tert-Butylbenzene	11.713	119	133092	19.127	ug/l	93
84) 1,2,4-Trimethylbenzene	11.750	105	127981	19.554	ug/l	95
85) sec-Butylbenzene	11.890	105	154825	19.408	ug/l	97
86) p-Isopropyltoluene	12.006	119	133708	19.111	ug/l	95
87) 1,3-Dichlorobenzene	11.969	146	76391	18.730	ug/l	95
88) 1,4-Dichlorobenzene	12.042	146	77986	18.549	ug/l	96
89) n-Butylbenzene	12.335	91	101325	18.294	ug/l	98
90) Hexachloroethane	12.536	117	21157	17.779	ug/l	66
91) 1,2-Dichlorobenzene	12.335	146	79398	19.111	ug/l	95
92) 1,2-Dibromo-3-Chloropr...	12.945	75	11823	21.314	ug/l	67
93) 1,2,4-Trichlorobenzene	13.591	180	50070	18.287	ug/l	95
94) Hexachlorobutadiene	13.725	225	23905	17.622	ug/l	99
95) Naphthalene	13.774	128	162807	20.150	ug/l	99
96) 1,2,3-Trichlorobenzene	13.963	180	52845	18.753	ug/l	96

06/12/2024

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061124\
Data File : VX041824.D
Acq On : 11 Jun 2024 12:46
Operator : JC/MD
Sample : VX0611MBSD01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 2 Sample Multiplier: 1

Quant Time: Jun 12 05:48:38 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
Quant Title : SW846 8260
QLast Update : Thu May 30 07:36:24 2024
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VX0611MBSD01

Manual Integrations
APPROVED


Reviewed By :Semsettin
Yesilyurt

06/12/2024
Supervised By :Mahesh
Dadoda

06/12/2024


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed						



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46

Instrument :
MSVOA_X
ClientSampleId :
VX0611MBSD01






Tgt Ion:168 Resp: 17335
Ion Ratio Lower Upper
168 100
99 51.2 56.6 85.0

Manual Integrations
APPROVED

Reviewed By :Semsettin
Yesilyurt




06/12/2024
Supervised By :Mahesh
Dadoda


06/12/2024



#2
Dichlorodifluoromethane
Concen: 18.490 ug/l
RT: 1.166 min Scan# 13
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46


Tgt Ion: 85 Resp: 28165
Ion Ratio Lower Upper
85 100
87 30.6 16.7 50.1





#3
Chloromethane
Concen: 18.835 ug/l
RT: 1.300 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46

Instrument :
MSVOA_X
ClientSampleId :
VX0611MBSD01





Tgt Ion: 50 Resp: 3130
Ion Ratio Lower Upper
50 100
52 31.6 25.5 38.3

Manual Integrations
APPROVED

Reviewed By :Semsettin
Yesilyurt




06/12/2024
Supervised By :Mahesh
Dadoda


06/12/2024



#4
Vinyl Chloride
Concen: 18.409 ug/l
RT: 1.373 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46


Tgt Ion: 62 Resp: 31496
Ion Ratio Lower Upper
62 100
64 32.3 25.4 38.0





#5
Bromomethane
Concen: 18.217 ug/l
RT: 1.599 min Scan# 84
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46

Instrument :
MSVOA_X
ClientSampleId :
VX0611MBSD01





Tgt Ion: 94 Resp: 21700
Ion Ratio Lower Upper
94 100
96 94.6 74.6 111.8

Manual Integrations
APPROVED

Reviewed By :Semsettin
Yesilyurt




06/12/2024
Supervised By :Mahesh
Dadoda

06/12/2024



#6
Chloroethane
Concen: 19.340 ug/l
RT: 1.666 min Scan# 95
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46


Tgt Ion: 64 Resp: 19999
Ion Ratio Lower Upper
64 100
66 31.4 25.1 37.7





#7
Trichlorofluoromethane
Concen: 20.214 ug/l
RT: 1.867 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46

Instrument :
MSVOA_X
ClientSampleId :
VX0611MBSD01





Tgt Ion:101 Resp: 56479
Ion Ratio Lower Upper
101 100
103 66.4 51.4 77.2

Manual Integrations
APPROVED

Reviewed By :Semsettin
Yesilyurt




06/12/2024
Supervised By :Mahesh
Dadoda


06/12/2024



#8
Diethyl Ether
Concen: 23.088 ug/l
RT: 2.136 min Scan# 172
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46


Tgt Ion: 74 Resp: 24860
Ion Ratio Lower Upper
74 100
45 100.2 53.3 159.8





#9
1,1,2-Trichlorotrifluoroethane
Concen: 21.346 ug/l
RT: 2.318 min Scan# 202
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46

Instrument :
MSVOA_X
ClientSampleId :
VX0611MBSD01






Tgt Ion:101 Resp: 33792
Ion Ratio Lower Upper
101 100
85 41.8 37.0 55.6
151 85.6 58.6 87.8

Manual Integrations
APPROVED

Reviewed By :Semsettin
Yesilyurt




06/12/2024
Supervised By :Mahesh
Dadoda


06/12/2024



#10
Methyl Iodide
Concen: 18.851 ug/l
RT: 2.446 min Scan# 223
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46


Tgt Ion:142 Resp: 37214
Ion Ratio Lower Upper
142 100
127 45.1 41.6 62.4
141 15.0 11.6 17.4





#11
Tert butyl alcohol
Concen: 109.524 ug/l
RT: 3.001 min Scan# 315
Delta R.T. -0.006 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46

Instrument :
MSVOA_X
ClientSampleId :
VX0611MBSD01






Tgt Ion: 59 Resp: 41193
Ion Ratio Lower Upper
59 100
57 10.4 6.0 9.0

Manual Integrations
APPROVED

Reviewed By :Semsettin
Yesilyurt




06/12/2024
Supervised By :Mahesh
Dadoda


06/12/2024



#12
1,1-Dichloroethene
Concen: 21.916 ug/l
RT: 2.312 min Scan# 201
Delta R.T. 0.006 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46


Tgt Ion: 96 Resp: 33867
Ion Ratio Lower Upper
96 100
61 153.2 144.6 216.8
98 62.2 50.4 75.6





#13
Acrolein
Concen: 114.480 ug/l
RT: 2.239 min Scan# 189
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46

Instrument :
MSVOA_X
ClientSampleId :
VX0611MBSD01






Tgt Ion: 56 Resp: 40090
Ion Ratio Lower Upper
56 100
55 70.5 57.1 85.7

Manual Integrations
APPROVED

Reviewed By :Semsettin
Yesilyurt




06/12/2024
Supervised By :Mahesh
Dadoda


06/12/2024



#14
Allyl chloride
Concen: 20.057 ug/l
RT: 2.660 min Scan# 258
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46


Tgt Ion: 41 Resp: 50988
Ion Ratio Lower Upper
41 100
39 69.3 55.8 83.6
76 31.2 24.3 36.5





#15
Acrylonitrile
Concen: 109.330 ug/l
RT: 3.068 min Scan# 31
Delta R.T. -0.006 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46

Instrument :
MSVOA_X
ClientSampleId :
VX0611MBSD01






Tgt Ion: 53 Resp: 100844
Ion Ratio Lower Upper
53 100
52 84.4 66.0 99.0
51 36.8 29.8 44.6

Manual Integrations
APPROVED

Reviewed By :Semsettin
Yesilyurt




06/12/2024
Supervised By :Mahesh
Dadoda


06/12/2024



#16
Acetone
Concen: 98.696 ug/l
RT: 2.392 min Scan# 214
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46

Tgt Ion: 43 Resp: 93693
Ion Ratio Lower Upper
43 100
58 29.2 21.9 32.9





#17

Carbon Disulfide

Concen: 15.903 ug/l

RT: 2.501 min Scan# 21

Delta R.T. -0.000 min

Lab File: VX041824.D

Acq: 11 Jun 2024 12:46

Instrument :

MSVOA_X

ClientSampleId :

VX0611MBSD01

Tgt Ion: 76 Resp: 5436

Ion Ratio Lower Upper

76 100

78 9.7 7.0 10.6

Manual Integrations





APPROVED

Reviewed By :Semsettin
Yesilyurt

06/12/2024

Supervised By :Mahesh
Dadoda

06/12/2024



#18

Methyl Acetate

Concen: 24.642 ug/l

RT: 2.709 min Scan# 266

Delta R.T. -0.000 min

Lab File: VX041824.D




Acq: 11 Jun 2024 12:46


Tgt Ion: 43 Resp: 50025

Ion Ratio Lower Upper

43 100


74 22.5 17.8 26.8





#19
Methyl tert-butyl Ether
Concen: 20.434 ug/l
RT: 3.117 min Scan# 311
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46

Instrument :
MSVOA_X
ClientSampleId :
VX0611MBSD01






Tgt Ion: 73 Resp: 10564
Ion Ratio Lower Upper
73 100
57 22.3 17.3 25.9

Manual Integrations
APPROVED

Reviewed By :Semsettin
Yesilyurt




06/12/2024
Supervised By :Mahesh
Dadoda


06/12/2024



#20
Methylene Chloride
Concen: 18.653 ug/l
RT: 2.788 min Scan# 279
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46


Tgt Ion: 84 Resp: 35491
Ion Ratio Lower Upper
84 100
49 125.4 103.6 155.4
51 38.4 33.8 50.6
86 67.7 49.4 74.0





#21
trans-1,2-Dichloroethene
Concen: 18.518 ug/l
RT: 3.087 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46

Instrument :
MSVOA_X
ClientSampleId :
VX0611MBSD01






Tgt Ion: 96 Resp: 29690
Ion Ratio Lower Upper
96 100
61 136.8 119.2 178.8
98 67.3 48.7 73.1

Manual Integrations
APPROVED

Reviewed By :Semsettin
Yesilyurt




06/12/2024
Supervised By :Mahesh
Dadoda


06/12/2024



#22
Diisopropyl ether
Concen: 21.131 ug/l
RT: 3.763 min Scan# 439
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46

Tgt Ion: 45 Resp: 109785
Ion Ratio Lower Upper
45 100
43 81.8 69.7 104.5
87 27.9 20.8 31.2
59 11.6 9.0 13.4





#23
Vinyl Acetate
Concen: 103.917 ug/l
RT: 3.721 min Scan# 41
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46

Instrument :
MSVOA_X
ClientSampleId :
VX0611MBSD01





Tgt Ion: 43 Resp: 48502
Ion Ratio Lower Upper
43 100
86 10.7 7.8 11.8

Manual Integrations
APPROVED

Reviewed By :Semsettin
Yesilyurt




06/12/2024
Supervised By :Mahesh
Dadoda


06/12/2024



#24
1,1-Dichloroethane
Concen: 19.995 ug/l
RT: 3.605 min Scan# 413
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46

Tgt Ion: 63 Resp: 58740
Ion Ratio Lower Upper
63 100
98 7.2 3.3 9.8
100 5.0 2.1 6.5





#25
2-Butanone
Concen: 107.854 ug/l
RT: 4.568 min Scan# 51
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46

Instrument :
MSVOA_X
ClientSampleId :
VX0611MBSD01





Tgt Ion: 43 Resp: 142519
Ion Ratio Lower Upper
43 100
72 25.3 19.4 29.2

Manual Integrations
APPROVED

Reviewed By :Semsettin
Yesilyurt




06/12/2024
Supervised By :Mahesh
Dadoda


06/12/2024



#26
2,2-Dichloropropane
Concen: 17.332 ug/l
RT: 4.464 min Scan# 554
Delta R.T. -0.006 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46

Tgt Ion: 77 Resp: 41776
Ion Ratio Lower Upper
77 100
97 22.6 10.8 32.3





#27

cis-1,2-Dichloroethene

Concen: 19.762 ug/l

RT: 4.489 min Scan# 51

Delta R.T. 0.006 min

Lab File: VX041824.D

Acq: 11 Jun 2024 12:46

Instrument :

MSVOA_X

ClientSampleId :

VX0611MBSD01

Tgt Ion: 96 Resp: 3787

Ion	Ratio	Lower	Upper
96	100		
61	142.8	0.0	312.2
98	64.4	0.0	129.0

Manual Integrations





APPROVED

Reviewed By :Semsettin
Yesilyurt

06/12/2024

Supervised By :Mahesh
Dadoda

06/12/2024



#28

Bromochloromethane

Concen: 18.811 ug/l

RT: 4.897 min Scan# 625




Delta R.T. -0.000 min


Lab File: VX041824.D

Acq: 11 Jun 2024 12:46

Tgt Ion: 49 Resp: 23802

Ion	Ratio	Lower	Upper
49	100		
129	1.5	0.0	3.8
130	86.2	57.5	86.3





#29

Tetrahydrofuran

Concen: 110.360 ug/l

RT: 5.013 min Scan# 644

Delta R.T. -0.000 min

Lab File: VX041824.D

Acq: 11 Jun 2024 12:46

Instrument :

MSVOA_X

ClientSampleId :

VX0611MBSD01

Tgt Ion: 42 Resp: 91968

Ion	Ratio	Lower	Upper
42	100		
72	44.4	34.2	51.4
71	40.9	32.2	48.2





Manual Integrations
APPROVED

Reviewed By :Semsettin
Yesilyurt

06/12/2024

Supervised By :Mahesh
Dadoda

06/12/2024



#30

Chloroform

Concen: 20.302 ug/l

RT: 5.092 min Scan# 657




Delta R.T. -0.000 min


Lab File: VX041824.D

Acq: 11 Jun 2024 12:46

Tgt Ion: 83 Resp: 62620


Ion	Ratio	Lower	Upper
83	100		
85	67.5	50.3	75.5





#31
Cyclohexane
Concen: 17.533 ug/l
RT: 5.464 min Scan# 718
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46

Instrument :
MSVOA_X
ClientSampleId :
VX0611MBSD01






Tgt Ion: 56 Resp: 4401
Ion Ratio Lower Upper
56 100
69 31.4 23.4 35.2
84 92.0 65.9 98.9

Manual Integrations
APPROVED

Reviewed By :Semsettin
Yesilyurt




06/12/2024
Supervised By :Mahesh
Dadoda


06/12/2024



#32
1,1,1-Trichloroethane
Concen: 19.272 ug/l
RT: 5.379 min Scan# 704
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46

Tgt Ion: 97 Resp: 52090
Ion Ratio Lower Upper
97 100
99 63.8 51.1 76.7
61 43.4 36.4 54.6





#33
1,2-Dichloroethane-d4
Concen: 49.782 ug/l
RT: 5.952 min Scan# 798
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46

Instrument :
MSVOA_X
ClientSampleId :
VX0611MBSD01





Tgt Ion: 65 Resp: 10584
Ion Ratio Lower Upper
65 100
67 50.7 0.0 100.2

Manual Integrations
APPROVED

Reviewed By :Semsettin
Yesilyurt




06/12/2024
Supervised By :Mahesh
Dadoda


06/12/2024



#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.757 min Scan# 930
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46


Tgt Ion:114 Resp: 258690
Ion Ratio Lower Upper
114 100
63 20.2 0.0 45.0
88 15.1 0.0 36.6





#35
Dibromofluoromethane
Concen: 46.492 ug/l
RT: 5.385 min Scan# 705
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46

Instrument :
MSVOA_X
ClientSampleId :
VX0611MBSD01






Tgt Ion: 113 Resp: 83154
Ion Ratio Lower Upper
113 100
111 102.9 82.9 124.3
192 24.9 13.3 19.9

Manual Integrations
APPROVED

Reviewed By :Semsettin
Yesilyurt




06/12/2024
Supervised By :Mahesh
Dadoda


06/12/2024



#36
1,1-Dichloropropene
Concen: 17.208 ug/l
RT: 5.690 min Scan# 755
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46

Tgt Ion: 75 Resp: 37267
Ion Ratio Lower Upper
75 100
110 37.6 17.0 50.9
77 30.6 24.7 37.1





#37
Ethyl Acetate
Concen: 19.333 ug/l
RT: 4.720 min Scan# 596
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46

Instrument :
MSVOA_X
ClientSampleId :
VX0611MBSD01





Tgt Ion: 43 Resp: 52839
Ion Ratio Lower Upper
43 100
61 13.6 10.6 16.0
70 10.3 8.3 12.5

Manual Integrations
APPROVED

Reviewed By :Semsettin
Yesilyurt




06/12/2024
Supervised By :Mahesh
Dadoda


06/12/2024



#38
Carbon Tetrachloride
Concen: 17.464 ug/l
RT: 5.672 min Scan# 752
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46

Tgt Ion:117 Resp: 42969
Ion Ratio Lower Upper
117 100
119 96.3 78.2 117.2
121 29.2 24.4 36.6





#39

Methylcyclohexane

Concen: 17.164 ug/l

RT: 7.379 min Scan# 1032

Delta R.T. -0.000 min

Lab File: VX041824.D

Acq: 11 Jun 2024 12:46

Instrument :

MSVOA_X

ClientSampleId :

VX0611MBSD01

Tgt Ion: 83 Resp: 4566

Ion	Ratio	Lower	Upper
83	100		
55	80.1	64.6	97.0
98	52.4	36.7	55.1

Manual Integrations





APPROVED

Reviewed By :Semsettin
Yesilyurt

06/12/2024

Supervised By :Mahesh
Dadoda

06/12/2024



#40

Benzene

Concen: 19.078 ug/l

RT: 6.037 min Scan# 812




Delta R.T. -0.000 min


Lab File: VX041824.D

Acq: 11 Jun 2024 12:46

Tgt Ion: 78 Resp: 129142


Ion	Ratio	Lower	Upper
78	100		
77	23.3	19.3	28.9





#41
Methacrylonitrile
Concen: 21.689 ug/l
RT: 4.928 min Scan# 61
Delta R.T. 0.006 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46

Instrument :
MSVOA_X
ClientSampleId :
VX0611MBSD01






Tgt Ion: 41 Resp: 2899
Ion Ratio Lower Upper
41 100
39 55.1 45.2 67.8
67 72.4 52.6 78.8
52 31.0 22.7 34.1

Manual Integrations
APPROVED

Reviewed By :Semsettin
Yesilyurt




06/12/2024
Supervised By :Mahesh
Dadoda


06/12/2024



#42
1,2-Dichloroethane
Concen: 20.052 ug/l
RT: 6.086 min Scan# 820
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46


Tgt Ion: 62 Resp: 50672
Ion Ratio Lower Upper
62 100
98 9.1 0.0 16.0





#43
Isopropyl Acetate
Concen: 20.109 ug/l
RT: 6.342 min Scan# 80
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46

Instrument :
MSVOA_X
ClientSampleId :
VX0611MBSD01






Tgt Ion: 43 Resp: 82978
Ion Ratio Lower Upper
43 100
61 19.3 15.2 22.8
87 13.1 9.8 14.6

Manual Integrations
APPROVED

Reviewed By :Semsettin
Yesilyurt




06/12/2024
Supervised By :Mahesh
Dadoda


06/12/2024



#44
Trichloroethene
Concen: 17.786 ug/l
RT: 7.122 min Scan# 990
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46

Tgt Ion:130 Resp: 32745
Ion Ratio Lower Upper
130 100
95 97.0 0.0 204.4





#45
1,2-Dichloropropane
Concen: 20.094 ug/l
RT: 7.427 min Scan# 1040
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46

Instrument :
MSVOA_X
ClientSampleId :
VX0611MBSD01





Tgt Ion: 63 Resp: 3234
Ion Ratio Lower Upper
63 100
65 30.9 24.3 36.5

Manual Integrations
APPROVED

Reviewed By :Semsettin
Yesilyurt




06/12/2024
Supervised By :Mahesh
Dadoda


06/12/2024



#46
Dibromomethane
Concen: 20.045 ug/l
RT: 7.580 min Scan# 1065
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46

Tgt Ion: 93 Resp: 24745
Ion Ratio Lower Upper
93 100
95 82.2 66.6 100.0
174 117.4 69.4 104.0#





#47
Bromodichloromethane
Concen: 19.296 ug/l
RT: 7.817 min Scan# 1105
Delta R.T. -0.006 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46

Instrument :
MSVOA_X
ClientSampleId :
VX0611MBSD01





Tgt Ion: 83 Resp: 4410
Ion Ratio Lower Upper
83 100
85 64.1 49.8 74.6
127 8.6 7.1 10.7

Manual Integrations
APPROVED

Reviewed By :Semsettin
Yesilyurt




06/12/2024
Supervised By :Mahesh
Dadoda


06/12/2024



#48
Methyl methacrylate
Concen: 20.684 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46


Tgt Ion: 41 Resp: 40989
Ion Ratio Lower Upper
41 100
69 81.0 59.2 88.8
39 60.0 47.7 71.5





#49
1,4-Dioxane
Concen: 423.437 ug/l
RT: 7.665 min Scan# 1080
Delta R.T. -0.006 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46

Instrument :
MSVOA_X
ClientSampleId :
VX0611MBSD01






Tgt Ion: 88 Resp: 18334
Ion Ratio Lower Upper
88 100
43 35.4 0.0 0.0
58 69.1 60.1 90.1

Manual Integrations
APPROVED

Reviewed By :Semsettin
Yesilyurt




06/12/2024
Supervised By :Mahesh
Dadoda


06/12/2024



#50
Toluene-d8
Concen: 45.606 ug/l
RT: 8.647 min Scan# 1240
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46


Tgt Ion: 98 Resp: 285015
Ion Ratio Lower Upper
98 100
100 66.3 53.1 79.7





#51
4-Methyl-2-Pentanone
Concen: 109.000 ug/l
RT: 8.573 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46

Instrument :
MSVOA_X
ClientSampleId :
VX0611MBSD01






Tgt Ion: 43 Resp: 29166
Ion Ratio Lower Upper
43 100
58 37.4 30.4 45.6

Manual Integrations
APPROVED

Reviewed By :Semsettin
Yesilyurt




06/12/2024
Supervised By :Mahesh
Dadoda


06/12/2024



#52
Toluene
Concen: 18.966 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. -0.006 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46


Tgt Ion: 92 Resp: 81476
Ion Ratio Lower Upper
92 100
91 169.4 134.2 201.2





#53
t-1,3-Dichloropropene
Concen: 18.490 ug/l
RT: 8.976 min Scan# 1194
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46



Instrument :
MSVOA_X
ClientSampleId :
VX0611MBSD01



Tgt Ion: 75 Resp: 4160
Ion Ratio Lower Upper
75 100
77 31.6 25.7 38.5


Manual Integrations
APPROVED

Reviewed By :Semsettin
Yesilyurt






06/12/2024
Supervised By :Mahesh
Dadoda


06/12/2024



#54
cis-1,3-Dichloropropene
Concen: 18.814 ug/l
RT: 8.366 min Scan# 1194
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46

Tgt Ion: 75 Resp: 46828
Ion Ratio Lower Upper
75 100
77 30.0 25.3 37.9
39 53.3 46.7 70.1






#55
1,1,2-Trichloroethane
Concen: 20.227 ug/l
RT: 9.153 min Scan# 1323
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46

Instrument :

MSVOA_X

ClientSampleId :

VX0611MBSD01



Tgt Ion: 97 Resp: 35028

Ion	Ratio	Lower	Upper
97	100		
83	83.0	65.8	98.8
85	53.2	41.6	62.4
99	63.2	47.8	71.8




Manual Integrations
APPROVED

Reviewed By :Semsettin
Yesilyurt

06/12/2024

Supervised By :Mahesh
Dadoda




06/12/2024




#56
Ethyl methacrylate
Concen: 20.264 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46

Tgt Ion: 69 Resp: 53562

Ion	Ratio	Lower	Upper
69	100		
41	72.5	56.8	85.2
39	40.7	32.0	48.0





#57
1,3-Dichloropropane
Concen: 20.289 ug/l
RT: 9.305 min Scan# 1173
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46

Instrument :
MSVOA_X
ClientSampleId :
VX0611MBSD01





Tgt Ion: 76 Resp: 5833
Ion Ratio Lower Upper
76 100
78 31.4 26.1 39.1

Manual Integrations
APPROVED

Reviewed By :Semsettin
Yesilyurt




06/12/2024
Supervised By :Mahesh
Dadoda


06/12/2024



#58
2-Chloroethyl Vinyl ether
Concen: 102.536 ug/l
RT: 8.238 min Scan# 1173
Delta R.T. -0.006 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46

Tgt Ion: 63 Resp: 127598
Ion Ratio Lower Upper
63 100
106 30.3 22.5 33.7





#59
2-Hexanone
Concen: 107.831 ug/l
RT: 9.433 min Scan# 1369
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46

Instrument :
MSVOA_X
ClientSampleId :
VX0611MBSD01





Tgt Ion: 43 Resp: 21998
Ion Ratio Lower Upper
43 100
58 54.4 26.2 78.5

Manual Integrations
APPROVED

Reviewed By :Semsettin
Yesilyurt




06/12/2024
Supervised By :Mahesh
Dadoda


06/12/2024



#60
Dibromochloromethane
Concen: 18.836 ug/l
RT: 9.518 min Scan# 1383
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46


Tgt Ion:129 Resp: 36186
Ion Ratio Lower Upper
129 100
127 77.0 38.8 116.4





#61
1,2-Dibromoethane
Concen: 19.716 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46

Instrument :
MSVOA_X
ClientSampleId :
VX0611MBSD01






Tgt Ion:107 Resp: 3603
Ion Ratio Lower Upper
107 100
109 94.3 75.2 112.8

Manual Integrations
APPROVED

Reviewed By :Semsettin
Yesilyurt




06/12/2024
Supervised By :Mahesh
Dadoda


06/12/2024



#62
4-Bromofluorobenzene
Concen: 45.479 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46


Tgt Ion: 95 Resp: 104901
Ion Ratio Lower Upper
95 100
174 94.6 0.0 135.6
176 92.4 0.0 131.0





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46

Instrument :
MSVOA_X
ClientSampleId :
VX0611MBSD01






Tgt Ion:117 Resp: 236100
Ion Ratio Lower Upper
117 100
82 51.9 46.2 69.4
119 32.3 25.5 38.3

Manual Integrations
APPROVED

Reviewed By :Semsettin
Yesilyurt




06/12/2024
Supervised By :Mahesh
Dadoda


06/12/2024



#64
Tetrachloroethene
Concen: 17.913 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. 0.006 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46

Tgt Ion:164 Resp: 31698
Ion Ratio Lower Upper
164 100
166 127.3 102.7 154.1
129 83.2 80.4 120.6
131 80.4 76.5 114.7





#65
Chlorobenzene
Concen: 18.452 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46

Instrument :
MSVOA_X
ClientSampleId :
VX0611MBSD01





Tgt Ion:112 Resp: 93490
Ion Ratio Lower Upper
112 100
114 31.6 25.7 38.5

Manual Integrations
APPROVED

Reviewed By :Semsettin
Yesilyurt




06/12/2024
Supervised By :Mahesh
Dadoda


06/12/2024



#66
1,1,1,2-Tetrachloroethane
Concen: 18.867 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. -0.006 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46


Tgt Ion:131 Resp: 33174
Ion Ratio Lower Upper
131 100
133 97.2 47.1 141.4
119 62.4 32.2 96.6





#67
Ethyl Benzene
Concen: 18.466 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46

Instrument :
MSVOA_X
ClientSampleId :
VX0611MBSD01





Tgt Ion: 91 Resp: 149120
Ion Ratio Lower Upper
91 100
106 32.2 24.9 37.3

Manual Integrations
APPROVED

Reviewed By :Semsettin
Yesilyurt




06/12/2024
Supervised By :Mahesh
Dadoda


06/12/2024



#68
m/p-Xylenes
Concen: 37.538 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46


Tgt Ion:106 Resp: 120591
Ion Ratio Lower Upper
106 100
91 195.6 165.5 248.3





#69
o-Xylene
Concen: 18.330 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46

Instrument :
MSVOA_X
ClientSampleId :
VX0611MBSD01






Tgt Ion:106 Resp: 58580
Ion Ratio Lower Upper
106 100
91 213.7 108.9 326.8

Manual Integrations
APPROVED

Reviewed By :Semsettin
Yesilyurt




06/12/2024
Supervised By :Mahesh
Dadoda


06/12/2024



#70
Styrene
Concen: 18.874 ug/l
RT: 10.652 min Scan# 1569
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46

Tgt Ion:104 Resp: 99626
Ion Ratio Lower Upper
104 100
78 49.6 43.6 65.4
103 55.0 43.6 65.4





#71

Bromoform

Concen: 18.848 ug/l

RT: 10.799 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX041824.D

Acq: 11 Jun 2024 12:46

Instrument :

MSVOA_X

ClientSampleId :

VX0611MBSD01

Tgt Ion:173	Resp:	2886
Ion Ratio	Lower	Upper
173	100	
175	48.3	24.1 72.4
254	0.0	0.0 0.0





Manual Integrations
APPROVED

Reviewed By :Semsettin
Yesilyurt

06/12/2024

Supervised By :Mahesh
Dadoda

06/12/2024



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 12.018 min Scan# 1793


Delta R.T. -0.006 min

Lab File: VX041824.D

Acq: 11 Jun 2024 12:46

Tgt Ion:152	Resp:	124729
Ion Ratio	Lower	Upper
152	100	
115	61.9	43.3 129.9
150	161.5	0.0 345.0





#73
Isopropylbenzene
Concen: 19.687 ug/l
RT: 10.963 min Scan# 1620
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46

Instrument :
MSVOA_X
ClientSampleId :
VX0611MBSD01





Tgt Ion:105 Resp: 152319
Ion Ratio Lower Upper
105 100
120 26.4 12.9 38.6

Manual Integrations
APPROVED

Reviewed By :Semsettin
Yesilyurt




06/12/2024
Supervised By :Mahesh
Dadoda


06/12/2024



#74
N-ethyl acetate
Concen: 20.864 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46

Tgt Ion: 43 Resp: 71506
Ion Ratio Lower Upper
43 100
70 42.9 32.9 49.3
55 25.1 19.3 28.9
61 23.9 18.8 28.2





#75
1,1,2,2-Tetrachloroethane
Concen: 21.102 ug/l
RT: 11.213 min Scan# 1661
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46

Instrument :
MSVOA_X
ClientSampleId :
VX0611MBSD01





Tgt Ion: 83 Resp: 5794
Ion Ratio Lower Upper
83 100
131 10.7 5.3 15.8
85 64.1 32.0 96.2

Manual Integrations
APPROVED

Reviewed By :Semsettin
Yesilyurt




06/12/2024
Supervised By :Mahesh
Dadoda


06/12/2024



#76
1,2,3-Trichloropropane
Concen: 21.395 ug/l m
RT: 11.238 min Scan# 1665
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46

Tgt Ion: 75 Resp: 46561
Ion Ratio Lower Upper
75 100
77 42.2 19.1 57.5





#77

Bromobenzene

Concen: 19.582 ug/l

RT: 11.195 min Scan# 1658

Delta R.T. -0.000 min

Lab File: VX041824.D

Acq: 11 Jun 2024 12:46

Instrument :

MSVOA_X

ClientSampleId :

VX0611MBSD01

Tgt Ion:156 Resp: 44720

Ion	Ratio	Lower	Upper
156	100		
77	136.7	88.4	265.3
158	95.8	48.0	144.0





Manual Integrations
APPROVED

Reviewed By :Semsettin
Yesilyurt

06/12/2024

Supervised By :Mahesh
Dadoda

06/12/2024



#78

n-propylbenzene

Concen: 19.239 ug/l

RT: 11.305 min Scan# 1676




Delta R.T. -0.000 min


Lab File: VX041824.D

Acq: 11 Jun 2024 12:46

Tgt Ion: 91 Resp: 163447

Ion	Ratio	Lower	Upper
91	100		
120	24.3	11.5	34.4





#79
2-Chlorotoluene
Concen: 19.344 ug/l
RT: 11.366 min Scan# 1100
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46

Instrument :
MSVOA_X
ClientSampleId :
VX0611MBSD01





Tgt Ion: 91 Resp: 10562
Ion Ratio Lower Upper
91 100
126 35.6 16.3 48.8

Manual Integrations
APPROVED

Reviewed By :Semsettin
Yesilyurt




06/12/2024
Supervised By :Mahesh
Dadoda


06/12/2024



#80
1,3,5-Trimethylbenzene
Concen: 19.570 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46


Tgt Ion:105 Resp: 126418
Ion Ratio Lower Upper
105 100
120 50.0 24.1 72.3





#81
trans-1,4-Dichloro-2-butene
Concen: 18.015 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46

Instrument :
MSVOA_X
ClientSampleId :
VX0611MBSD01






Tgt Ion: 75 Resp: 1438
Ion Ratio Lower Upper
75 100
53 117.3 90.5 135.7
89 45.8 36.9 55.3

Manual Integrations
APPROVED

Reviewed By :Semsettin
Yesilyurt




06/12/2024
Supervised By :Mahesh
Dadoda


06/12/2024



#82
4-Chlorotoluene
Concen: 19.180 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46


Tgt Ion: 91 Resp: 119149
Ion Ratio Lower Upper
91 100
126 31.2 14.2 42.6





#83
tert-Butylbenzene
Concen: 19.127 ug/l
RT: 11.713 min Scan# 1743
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46

Instrument :
MSVOA_X
ClientSampleId :
VX0611MBSD01






Tgt Ion:119 Resp: 133092
Ion Ratio Lower Upper
119 100
91 55.9 31.2 93.6
134 24.9 11.7 35.0

Manual Integrations
APPROVED

Reviewed By :Semsettin
Yesilyurt




06/12/2024
Supervised By :Mahesh
Dadoda


06/12/2024



#84
1,2,4-Trimethylbenzene
Concen: 19.554 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46


Tgt Ion:105 Resp: 127981
Ion Ratio Lower Upper
105 100
120 45.4 21.2 63.6





#85
sec-Butylbenzene
Concen: 19.408 ug/l
RT: 11.890 min Scan# 1772
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46

Instrument :
MSVOA_X
ClientSampleId :
VX0611MBSD01





Tgt Ion:105 Resp: 154825
Ion Ratio Lower Upper
105 100
134 20.9 9.8 29.4

Manual Integrations
APPROVED

Reviewed By :Semsettin
Yesilyurt




06/12/2024
Supervised By :Mahesh
Dadoda


06/12/2024



#86
p-Isopropyltoluene
Concen: 19.111 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46

Tgt Ion:119 Resp: 133708
Ion Ratio Lower Upper
119 100
134 27.2 12.8 38.3
91 22.9 13.1 39.3





#87

1,3-Dichlorobenzene

Concen: 18.730 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. -0.000 min

Lab File: VX041824.D


Acq: 11 Jun 2024 12:46

Instrument :

MSVOA_X

ClientSampleId :

VX0611MBSD01



Tgt Ion:146 Resp: 7639

Ion	Ratio	Lower	Upper
146	100		
111	37.0	22.1	66.1
148	64.7	31.8	95.3



Manual Integrations
APPROVED

Reviewed By :Semsettin
Yesilyurt

06/12/2024

Supervised By :Mahesh
Dadoda

06/12/2024



#88

1,4-Dichlorobenzene

Concen: 18.549 ug/l

RT: 12.042 min Scan# 1797




Delta R.T. -0.000 min


Lab File: VX041824.D

Acq: 11 Jun 2024 12:46

Tgt Ion:146 Resp: 77986

Ion	Ratio	Lower	Upper
146	100		
111	37.8	21.3	63.7
148	64.7	31.8	95.4





#89

n-Butylbenzene

Concen: 18.294 ug/l

RT: 12.335 min Scan# 1845

Delta R.T. 0.006 min

Lab File: VX041824.D

Acq: 11 Jun 2024 12:46

Instrument :

MSVOA_X

ClientSampleId :

VX0611MBSD01

Tgt Ion: 91 Resp: 10132

Ion	Ratio	Lower	Upper
91	100		
92	53.9	27.4	82.0
134	28.0	12.6	37.8

Manual Integrations





APPROVED

Reviewed By :Semsettin
Yesilyurt

06/12/2024

Supervised By :Mahesh
Dadoda

06/12/2024



#90

Hexachloroethane

Concen: 17.779 ug/l

RT: 12.536 min Scan# 1878




Delta R.T. -0.000 min


Lab File: VX041824.D

Acq: 11 Jun 2024 12:46

Tgt Ion:117 Resp: 21157


Ion	Ratio	Lower	Upper
117	100		
201	92.7	32.9	98.6





#91
1,2-Dichlorobenzene
Concen: 19.111 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46

Instrument : MSVOA_X
ClientSampleId : VX0611MBSD01






Tgt Ion:146 Resp: 79398
Ion Ratio Lower Upper
146 100
111 39.0 22.4 67.0
148 64.8 31.6 95.0

Manual Integrations
APPROVED


Reviewed By :Semsettin
Yesilyurt

06/12/2024
Supervised By :Mahesh
Dadoda



06/12/2024




#92
1,2-Dibromo-3-Chloropropane
Concen: 21.314 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46




Tgt Ion: 75 Resp: 11823
Ion Ratio Lower Upper
75 100
155 97.7 36.1 108.5
157 122.1 44.9 134.7





#93
1,2,4-Trichlorobenzene
Concen: 18.287 ug/l
RT: 13.591 min Scan# 2051
Delta R.T. 0.006 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46

Instrument :
MSVOA_X
ClientSampleId :
VX0611MBSD01






Tgt Ion:180 Resp: 50070
Ion Ratio Lower Upper
180 100
182 94.1 48.3 144.8
145 28.0 18.0 54.0

Manual Integrations
APPROVED

Reviewed By :Semsettin
Yesilyurt




06/12/2024
Supervised By :Mahesh
Dadoda


06/12/2024



#94
Hexachlorobutadiene
Concen: 17.622 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. -0.000 min
Lab File: VX041824.D
Acq: 11 Jun 2024 12:46

Tgt Ion:225 Resp: 23905
Ion Ratio Lower Upper
225 100
223 63.0 31.1 93.2
227 64.2 32.5 97.5





#95

Naphthalene

Concen: 20.150 ug/l

RT: 13.774 min Scan# 2081

Delta R.T. -0.000 min

Lab File: VX041824.D

Acq: 11 Jun 2024 12:46

Instrument :

MSVOA_X

ClientSampleId :

VX0611MBSD01

Tgt Ion:128 Resp: 16280

Ion	Ratio	Lower	Upper
128	100		
127	12.5	10.3	15.5
129	10.8	8.8	13.2

Manual Integrations





APPROVED

Reviewed By :Semsettin
Yesilyurt

06/12/2024

Supervised By :Mahesh
Dadoda

06/12/2024



#96

1,2,3-Trichlorobenzene

Concen: 18.753 ug/l

RT: 13.963 min Scan# 2112

Delta R.T. -0.000 min

Lab File: VX041824.D

Acq: 11 Jun 2024 12:46

Tgt Ion:180 Resp: 52845

Ion	Ratio	Lower	Upper
180	100		
182	95.8	46.9	140.6
145	30.4	17.8	53.5

