



BFB

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061124\
Data File : VX041819.D
Acq On : 11 Jun 2024 09:26
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
Title : SW846 8260
Last Update : Thu May 30 07:36:24 2024



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.1	7135	PASS
75	95	30	60	50.5	17945	PASS
95	95	100	100	100.0	35533	PASS
96	95	5	9	7.3	2599	PASS
173	174	0.00	2	0.4	128	PASS
174	95	50	100	91.6	32563	PASS
175	174	5	9	7.7	2513	PASS
176	174	95	101	96.5	31429	PASS
177	176	5	9	6.6	2088	PASS