

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061218\
 Data File : VX002469.D
 Acq On : 12 Jun 2018 16:24
 Operator : JC/MD
 Sample : J3402-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LF5-0073

Quant Time: Jun 13 10:59:58 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 05 03:22:35 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	199403	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	292552	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	260006	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	131635	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	160903	56.91	ug/l	0.00
Spiked Amount	50.000		Recovery	=	113.82%	
35) Dibromofluoromethane	5.51	113	121401	52.77	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.54%	
50) Toluene-d8	8.72	98	467487	53.01	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.02%	
62) 4-Bromofluorobenzene	11.14	95	151626	44.59	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.18%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061218\
Data File : VX002469.D
Acq On : 12 Jun 2018 16:24
Operator : JC/MD
Sample : J3402-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LF5-0073

Quant Time: Jun 13 10:59:58 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X060418W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 05 03:22:35 2018
Response via : Initial Calibration

