

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061218\
 Data File : VX002490.D
 Acq On : 13 Jun 2018 02:38
 Operator : JC/MD
 Sample : J3444-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SB-51(0-5)

Quant Time: Jun 13 03:58:55 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 05 03:22:35 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	214249	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	318035	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	302721	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	175780	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	157359	51.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.92%	
35) Dibromofluoromethane	5.51	113	115579	46.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.42%	
50) Toluene-d8	8.72	98	465950	48.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.22%	
62) 4-Bromofluorobenzene	11.14	95	163914	44.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.68%	
Target Compounds						
18) Methyl Acetate	2.78	43	21252	4.466	ug/l	95
25) 2-Butanone	4.71	43	4957	2.360	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061218\
Data File : VX002490.D
Acq On : 13 Jun 2018 02:38
Operator : JC/MD
Sample : J3444-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SB-51(0-5)

Quant Time: Jun 13 03:58:55 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X060418W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 05 03:22:35 2018
Response via : Initial Calibration

