

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061218\
 Data File : VX002463.D
 Acq On : 12 Jun 2018 13:44
 Operator : JC/MD
 Sample : J3329-11
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-29(0-5)

Quant Time: Jun 12 15:09:39 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 05 03:22:35 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	213188	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	313279	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	297169	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	170912	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	155293	51.01	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.02%	
35) Dibromofluoromethane	5.51	113	115881	47.04	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.08%	
50) Toluene-d8	8.72	98	453984	48.08	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.16%	
62) 4-Bromofluorobenzene	11.14	95	161642	44.39	ug/l	0.00
Spiked Amount	50.000		Recovery	=	88.78%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.67 min Scan# 833

Delta R.T. -0.01 min

Lab File: VX002463.D

Acq: 12 Jun 2018 13:44

Instrument :

MSVOA_X

ClientSampled :


MH-29(0-5)

Tot Ion:168 Resp: 213188

Ion Ratio Lower Upper

168 100

99 52.0 39.3 58.9



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.67

80000

60000



40000

20000

0

Time-->

5.60 5.70 5.80



#33

1,2-Dichloroethane-d4

Concen: 51.009 ug/l

RT: 6.07 min Scan# 900

Delta R.T. -0.00 min

Lab File: VX002463.D


Acq: 12 Jun 2018 13:44

Tgt Ion: 65 Resp: 155293

Ion Ratio Lower Upper

65 100

67 51.3 0.0 102.6



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0

6.07

60000


40000


20000

0

Time-->

6.00 6.10 6.20





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.87 min Scan# 1030

Delta R.T. -0.00 min

Lab File: VX002463.D

Acq: 12 Jun 2018 13:44

Instrument :

MSVOA_X

ClientSampleId :

MH-29(0-5)


Tot Ion:114 Resp: 313279

Ion Ratio Lower Upper

114 100

63 21.6 0.0 41.4




88 15.0 0.0 30.0



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 47.036 ug/l

RT: 5.51 min Scan# 807

Delta R.T. -0.00 min

Lab File: VX002463.D

Acq: 12 Jun 2018 13:44


Tgt Ion:113 Resp: 115881

Ion Ratio Lower Upper

113 100

111 102.9 81.4 122.0



192 22.6 19.1 28.7




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V









#50
Toluene-d8
Concen: 48.076 ug/l
RT: 8.72 min Scan# 1334
Delta R.T. -0.00 min
Lab File: VX002463.D
Acq: 12 Jun 2018 13:44




Instrument :
MSVOA_X
ClientSampleId :
MH-29(0-5)

Tot Ion: 98 Resp: 453984
Ion Ratio Lower Upper
98 100
100 63.4 50.9 76.3



#62
4-Bromofluorobenzene
Concen: 44.390 ug/l
RT: 11.14 min Scan# 1731
Delta R.T. -0.00 min
Lab File: VX002463.D
Acq: 12 Jun 2018 13:44

Tgt Ion: 95 Resp: 161642
Ion Ratio Lower Upper
95 100
174 94.1 0.0 187.4
176 91.5 0.0 181.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.12 min Scan# 1563
Delta R.T. -0.00 min
Lab File: VX002463.D
Acq: 12 Jun 2018 13:44

Instrument :
MSVOA_X
ClientSampled :
MH-29(0-5)

Tot Ion:117 Resp: 297169
Ion Ratio Lower Upper
117 100
82 55.2 45.0 67.4
119 33.0 25.8 38.6




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VX0
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.09 min Scan# 1886
Delta R.T. -0.00 min
Lab File: VX002463.D
Acq: 12 Jun 2018 13:44

Tgt Ion:152 Resp: 170912
Ion Ratio Lower Upper
152 100
115 55.6 40.1 120.2
150 155.7 0.0 347.2



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

