

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061218\
 Data File : VX002477.D
 Acq On : 12 Jun 2018 19:58
 Operator : JC/MD
 Sample : J3402-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LF5-0070

Quant Time: Jun 13 02:28:09 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 05 03:22:35 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	192335	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	285549	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	263124	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	143251	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	154749	56.73	ug/l	0.00
Spiked Amount	50.000		Recovery	=	113.46%	
35) Dibromofluoromethane	5.51	113	115674	51.51	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.02%	
50) Toluene-d8	8.72	98	444493	51.64	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.28%	
62) 4-Bromofluorobenzene	11.14	95	151198	45.55	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.10%	


Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061218\
Data File : VX002477.D
Acq On : 12 Jun 2018 19:58
Operator : JC/MD
Sample : J3402-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LF5-0070

Quant Time: Jun 13 02:28:09 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X060418W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 05 03:22:35 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.67 min Scan# 834

Delta R.T. -0.00 min

Lab File: VX002477.D

Acq: 12 Jun 2018 19:58

Instrument :

MSVOA_X

ClientSampleId :


LF5-0070

Tot Ion:168 Resp: 192335

Ion Ratio Lower Upper

168 100

99 52.4 39.3 58.9



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.67

80000



60000

40000

20000

0

Time--> 5.50 5.60 5.70 5.80



#33

1,2-Dichloroethane-d4

Concen: 56.730 ug/l

RT: 6.07 min Scan# 900

Delta R.T. -0.00 min

Lab File: VX002477.D


Acq: 12 Jun 2018 19:58

Tgt Ion: 65 Resp: 154749

Ion Ratio Lower Upper

65 100

67 51.0 0.0 102.6



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0

6.07


60000


40000

20000

0

Time--> 5.90 6.00 6.10 6.20 6.30





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.87 min Scan# 1030

Delta R.T. -0.00 min

Lab File: VX002477.D

Acq: 12 Jun 2018 19:58

Instrument :

MSVOA_X

ClientSampleId :

LF5-0070


Tot Ion:114 Resp: 285549

Ion Ratio Lower Upper

114 100

63 21.1 0.0 41.4




88 14.4 0.0 30.0



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 51.512 ug/l

RT: 5.51 min Scan# 807

Delta R.T. -0.00 min

Lab File: VX002477.D

Acq: 12 Jun 2018 19:58


Tgt Ion:113 Resp: 115674

Ion Ratio Lower Upper

113 100

111 101.4 81.4 122.0



192 22.3 19.1 28.7




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 51.642 ug/l

RT: 8.72 min Scan# 1334

Delta R.T. -0.00 min

Lab File: VX002477.D

Acq: 12 Jun 2018 19:58

Instrument :

MSVOA_X

ClientSampleId :


LF5-0070

Tot Ion: 98 Resp: 444493

Ion Ratio Lower Upper




98 100

100 63.2 50.9 76.3



Abundance Ion 98.00 (97.70 to 98.70): VX0

Ion 100.00 (99.70 to 100.70): VX



#62

4-Bromofluorobenzene

Concen: 45.555 ug/l

RT: 11.14 min Scan# 1731

Delta R.T. -0.00 min

Lab File: VX002477.D

Acq: 12 Jun 2018 19:58


Tgt Ion: 95 Resp: 151198

Ion Ratio Lower Upper

95 100

174 94.1 0.0 187.4



176 90.3 0.0 181.0



Abundance Ion 94.90 (94.60 to 95.60): VX0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.12 min Scan# 1563
Delta R.T. -0.00 min
Lab File: VX002477.D
Acq: 12 Jun 2018 19:58

Instrument :
MSVOA_X
ClientSampleId :
LF5-0070

Tot Ion:117 Resp: 263124
Ion Ratio Lower Upper
117 100
82 55.9 45.0 67.4
119 32.4 25.8 38.6




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VX0
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.09 min Scan# 1886
Delta R.T. -0.00 min
Lab File: VX002477.D
Acq: 12 Jun 2018 19:58

Tgt Ion:152 Resp: 143251
Ion Ratio Lower Upper
152 100
115 55.1 40.1 120.2
150 157.5 0.0 347.2



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

