

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061318\
 Data File : VX002535.D
 Acq On : 14 Jun 2018 04:01
 Operator : JC/MD
 Sample : J3383-03
 Misc : 6.10g/5mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Jun 14 04:57:01 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 05 03:22:35 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	260383	50.00	ug/l	-0.01
34) 1,4-Difluorobenzene	6.86	114	372910	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	352859	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	211092	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	170989	45.62	ug/l	0.00
Spiked Amount			Recovery	=	91.24%	
35) Dibromofluoromethane	5.51	113	125662	42.85	ug/l	0.00
Spiked Amount			Recovery	=	85.70%	
50) Toluene-d8	8.72	98	531744	47.31	ug/l	0.00
Spiked Amount			Recovery	=	94.62%	
62) 4-Bromofluorobenzene	11.15	95	196808	45.41	ug/l	0.00
Spiked Amount			Recovery	=	90.82%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) Chloromethane	1.32	50	363	Below Cal	#	84
5) Bromomethane	1.64	94	2088	Below Cal		97
8) Diethyl Ether	2.01	74	115	Below Cal		67
10) Methyl Iodide	2.49	142	1584	0.40	ug/l	94
11) Tert butyl alcohol	3.03	59	24235	13.00	ug/l	# 75
13) Acrolein	2.29	56	210	Below Cal		93
15) Acrylonitrile	3.16	53	109	Below Cal	#	17
16) Acetone	2.45	43	87	Below Cal	#	66
18) Methyl Acetate	2.80	43	3283	Below Cal		98
19) Methyl tert-butyl Ether	3.14	73	97	Below Cal	#	1
20) Methylene Chloride	2.85	84	866	Below Cal	#	80
22) Diisopropyl ether	3.87	45	106	Below Cal	#	36
28) Bromochloromethane	5.02	49	25	Below Cal	#	13
31) Cyclohexane	5.57	56	1928	0.41	ug/l	87
36) 1,1-Dichloropropene	5.67	75	18117	4.62	ug/l	# 51
37) Ethyl Acetate	4.75	43	3581	0.76	ug/l	# 71
38) Carbon Tetrachloride	5.66	117	24113	5.88	ug/l	# 15
39) Methylcyclohexane	7.46	83	3395	0.96	ug/l	98
67) Ethyl Benzene	10.26	91	4445	0.30	ug/l	97
68) m/p-Xylenes	10.37	106	1527	0.27	ug/l	89
76) 1,2,3-Trichloropropane	11.15	75	103298	22.30	ug/l	# 33
78) n-propylbenzene	11.37	91	3665	0.22	ug/l	97
81) trans-1,4-Dichloro-2-buten	11.15	75	103298	78.17	ug/l	# 6
84) 1,2,4-Trimethylbenzene	11.81	105	8741	0.67	ug/l	97
95) Naphthalene	13.84	128	3632	0.24	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061318\
Data File : VX002535.D
Acq On : 14 Jun 2018 04:01
Operator : JC/MD
Sample : J3383-03
Misc : 6.10q/5mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Jun 14 04:57:01 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X060418W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 05 03:22:35 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.00 ug/l

RT: 5.66 min Scan# 832

Delta R.T. -0.01 min

Lab File: VX002535.D

Acq: 14 Jun 2018 04:01

Instrument :

MSVOA_X


ClientSampleId :

Tot Ion:168 Resp: 260383

Ion Ratio Lower Upper




168 100

99 50.5 39.3 58.9



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



#3

Chloromethane

Concen: Below Cal

RT: 1.32 min Scan# 121

Delta R.T. 0.00 min

Lab File: VX002535.D


Acq: 14 Jun 2018 04:01

Tgt Ion: 50 Resp: 363

Ion Ratio Lower Upper



50 100

52 40.9 25.7 38.5#



Abundance Ion 49.90 (49.60 to 50.60): VX0

Ion 51.90 (51.60 to 52.60): VX0





#5

Bromomethane

Concen: Below Cal

RT: 1.64 min Scan# 172

Delta R.T. 0.00 min

Lab File: VX002535.D

Acq: 14 Jun 2018 04:01

Instrument :

MSVOA_X


ClientSampled :

Tot Ion: 94 Resp: 2088

Ion Ratio Lower Upper




94 100

96 96.5 74.9 112.3



Abundance Ion 93.90 (93.60 to 94.60): VX0

Ion 95.90 (95.60 to 96.60): VX0



#8

Diethyl Ether

Concen: Below Cal

RT: 2.01 min Scan# 233

Delta R.T. -0.18 min

Lab File: VX002535.D


Acq: 14 Jun 2018 04:01

Tgt Ion: 74 Resp: 115

Ion Ratio Lower Upper



74 100


45 140.9 53.1 159.4



Abundance Ion 74.00 (73.70 to 74.70): VX0

Ion 45.00 (44.70 to 45.70): VX0





#10

Methyl Iodide

Concen: 0.40 ug/l

RT: 2.49 min Scan# 312

Delta R.T. -0.02 min

Lab File: VX002535.D

Acq: 14 Jun 2018 04:01

Instrument :

MSVOA_X

ClientSampled :


Tot Ion:142 Resp: 1584

Ion Ratio Lower Upper

142 100

127 51.1 36.6 55.0




141 14.3 11.3 16.9



Abundance Ion 141.80 (141.50 to 142.50): V

Ion 126.80 (126.50 to 127.50): V

Ion 140.80 (140.50 to 141.50): V



#11

Tert butyl alcohol

Concen: 13.00 ug/l

RT: 3.03 min Scan# 400

Delta R.T. -0.05 min

Lab File: VX002535.D


Acq: 14 Jun 2018 04:01

Tgt Ion: 59 Resp: 24235

Ion Ratio Lower Upper



59 100


57 1.1 8.2 12.2#



Abundance Ion 59.00 (58.70 to 59.70): VX0

Ion 57.00 (56.70 to 57.70): VX0





#13

Acrolein

Concen: Below Cal

RT: 2.29 min Scan# 280

Delta R.T. -0.01 min

Lab File: VX002535.D

Acq: 14 Jun 2018 04:01

Instrument :

MSVOA_X


ClientSampled :

Tot Ion: 56 Resp: 210

Ion Ratio Lower Upper




56 100

55 65.2 57.0 85.4



Abundance Ion 56.00 (55.70 to 56.70): VX0

Ion 55.00 (54.70 to 55.70): VX0



#15

Acrylonitrile

Concen: Below Cal

RT: 3.16 min Scan# 422

Delta R.T. 0.01 min

Lab File: VX002535.D

Acq: 14 Jun 2018 04:01


Tgt Ion: 53 Resp: 109

Ion Ratio Lower Upper

53 100

52 0.0 66.7 100.1#



51 0.0 29.9 44.9#




Abundance Ion 53.00 (52.70 to 53.70): VX0

Ion 52.00 (51.70 to 52.70): VX0

Ion 51.00 (50.70 to 51.70): VX0





#16

Acetone

Concen: Below Cal

RT: 2.45 min Scan# 305

Delta R.T. -0.01 min

Lab File: VX002535.D

Acq: 14 Jun 2018 04:01

Instrument :

MSVOA_X


ClientSampleId :

Tot Ion: 43 Resp: 87

Ion Ratio Lower Upper




43 100

58 45.5 22.1 33.1#



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 58.00 (57.70 to 58.70): VX0



#18

Methyl Acetate

Concen: Below Cal

RT: 2.80 min Scan# 363

Delta R.T. 0.02 min

Lab File: VX002535.D


Acq: 14 Jun 2018 04:01

Tgt Ion: 43 Resp: 3283

Ion Ratio Lower Upper



43 100


74 20.0 16.7 25.1



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 74.00 (73.70 to 74.70): VX0









#19
Methyl tert-butyl Ether
Concen: Below Cal
RT: 3.14 min Scan# 418
Delta R.T. -0.07 min
Lab File: VX002535.D
Acq: 14 Jun 2018 04:01




Instrument :
MSVOA_X
ClientSampleId :


Tot Ion: 73 Resp: 97
Ion Ratio Lower Upper
73 100
57 864.2 17.7 26.5#



#20
Methylene Chloride
Concen: Below Cal
RT: 2.85 min Scan# 371
Delta R.T. -0.01 min
Lab File: VX002535.D
Acq: 14 Jun 2018 04:01

Tgt Ion: 84 Resp: 866
Ion Ratio Lower Upper
84 100
49 94.1 107.8 161.6#
51 40.0 32.8 49.2
86 67.3 52.5 78.7









#22
Diisopropyl ether
Concen: Below Cal
RT: 3.87 min Scan# 539
Delta R.T. -0.01 min
Lab File: VX002535.D
Acq: 14 Jun 2018 04:01




Instrument :
MSVOA_X
ClientSampleId :


Tot Ion: 45 Resp: 106
Ion Ratio Lower Upper
45 100
43 0.9 46.8 70.2#
87 0.0 20.2 30.4#
59 0.0 8.7 13.1#



#28
Bromochloromethane
Concen: Below Cal
RT: 5.02 min Scan# 727
Delta R.T. -0.01 min
Lab File: VX002535.D
Acq: 14 Jun 2018 04:01

Tgt Ion: 49 Resp: 25
Ion Ratio Lower Upper
49 100
129 0.0 0.0 4.2
130 0.0 61.2 91.8#









#31
Cyclohexane
Concen: 0.41 ug/l
RT: 5.57 min Scan# 818
Delta R.T. -0.01 min
Lab File: VX002535.D
Acq: 14 Jun 2018 04:01

Instrument :
MSVOA_X
ClientSampleId :

Tot Ion: 56 Resp: 1928
Ion Ratio Lower Upper
56 100
69 25.9 23.7 35.5
84 66.3 64.1 96.1




Abundance Ion 56.00 (55.70 to 56.70): VX0
Ion 69.00 (68.70 to 69.70): VX0
Ion 84.00 (83.70 to 84.70): VX0






#33
1,2-Dichloroethane-d4
Concen: 45.62 ug/l
RT: 6.07 min Scan# 899
Delta R.T. -0.01 min
Lab File: VX002535.D
Acq: 14 Jun 2018 04:01

Tgt Ion: 65 Resp: 170989
Ion Ratio Lower Upper
65 100
67 52.5 0.0 102.6



Abundance Ion 64.90 (64.60 to 65.60): VX0
Ion 66.90 (66.60 to 67.60): VX0





#34

1,4-Difluorobenzene

Concen: 50.00 ug/l

RT: 6.86 min Scan# 1029

Delta R.T. -0.01 min

Lab File: VX002535.D

Acq: 14 Jun 2018 04:01

Instrument :

MSVOA_X

ClientSampleId :


Tot Ion:114 Resp: 372910

Ion Ratio Lower Upper

114 100

63 21.0 0.0 41.4




88 15.8 0.0 30.0



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 42.85 ug/l

RT: 5.51 min Scan# 807

Delta R.T. -0.00 min

Lab File: VX002535.D

Acq: 14 Jun 2018 04:01


Tgt Ion:113 Resp: 125662

Ion Ratio Lower Upper

113 100

111 101.9 81.4 122.0



192 23.6 19.1 28.7




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V









#36
 1,1-Dichloropropene
 Concen: 4.62 ug/l
 RT: 5.67 min Scan# 833
 Delta R.T. -0.14 min
 Lab File: VX002535.D
 Acq: 14 Jun 2018 04:01

Instrument :
 MSVOA_X
 ClientSampleId :

Tot Ion: 75 Resp: 18117
 Ion Ratio Lower Upper
 75 100
 110 10.3 18.1 54.4#
 77 0.0 24.3 36.5#




Abundance Ion 74.90 (74.60 to 75.60): VX0
 Ion 109.90 (109.60 to 110.60): V
 Ion 76.90 (76.60 to 77.60): VX0






#37
 Ethyl Acetate
 Concen: 0.76 ug/l
 RT: 4.75 min Scan# 683
 Delta R.T. -0.12 min
 Lab File: VX002535.D
 Acq: 14 Jun 2018 04:01

Tgt Ion: 43 Resp: 3581
 Ion Ratio Lower Upper
 43 100
 61 0.0 10.4 15.6#
 70 1.1 7.6 11.4#



Abundance Ion 43.00 (42.70 to 43.70): VX0
 Ion 60.90 (60.60 to 61.60): VX0
 Ion 70.00 (69.70 to 70.70): VX0





#38

Carbon Tetrachloride

Concen: 5.88 ug/l

RT: 5.66 min Scan# 832

Delta R.T. -0.13 min

Lab File: VX002535.D

Acq: 14 Jun 2018 04:01

Instrument :

MSVOA_X

ClientSampled :


Tot Ion:117 Resp: 24113

Ion Ratio Lower Upper

117 100

119 4.7 77.4 116.0#




121 0.0 24.4 36.6#



Abundance Ion 116.80 (116.50 to 117.50): V

Ion 118.80 (118.50 to 119.50): V

Ion 120.80 (120.50 to 121.50): V



#39

Methylcyclohexane

Concen: 0.96 ug/l

RT: 7.46 min Scan# 1128

Delta R.T. -0.00 min

Lab File: VX002535.D

Acq: 14 Jun 2018 04:01


Tgt Ion: 83 Resp: 3395

Ion Ratio Lower Upper

83 100

55 79.0 65.4 98.0



98 46.9 37.4 56.0




Abundance Ion 83.00 (82.70 to 83.70): VX0

Ion 55.00 (54.70 to 55.70): VX0

Ion 98.00 (97.70 to 98.70): VX0









#50
Toluene-d8
Concen: 47.31 ug/l
RT: 8.72 min Scan# 1334
Delta R.T. -0.00 min
Lab File: VX002535.D
Acq: 14 Jun 2018 04:01




Instrument :
MSVOA_X
ClientSampleId :


Tot Ion: 98 Resp: 531744
Ion Ratio Lower Upper
98 100
100 63.6 50.9 76.3



#62
4-Bromofluorobenzene
Concen: 45.41 ug/l
RT: 11.15 min Scan# 1732
Delta R.T. 0.01 min
Lab File: VX002535.D
Acq: 14 Jun 2018 04:01

Tgt Ion: 95 Resp: 196808
Ion Ratio Lower Upper
95 100
174 95.8 0.0 187.4
176 93.1 0.0 181.0





#63

Chlorobenzene-d5

Concen: 50.00 ug/l

RT: 10.12 min Scan# 1564

Delta R.T. 0.01 min

Lab File: VX002535.D

Acq: 14 Jun 2018 04:01

Instrument :

MSVOA_X

ClientSampleId :


Tot Ion:117 Resp: 352859

Ion Ratio Lower Upper

117 100

82 53.4 45.0 67.4




119 32.6 25.8 38.6



Abundance Ion 116.90 (116.60 to 117.60): V

Ion 82.00 (81.70 to 82.70): VX0

Ion 118.90 (118.60 to 119.60): V



#67

Ethyl Benzene

Concen: 0.30 ug/l

RT: 10.26 min Scan# 1586

Delta R.T. -0.00 min

Lab File: VX002535.D


Acq: 14 Jun 2018 04:01

Tgt Ion: 91 Resp: 4445

Ion Ratio Lower Upper



91 100


106 30.0 25.5 38.3



Abundance Ion 91.00 (90.70 to 91.70): VX0

Ion 106.00 (105.70 to 106.70): V









#68
m/p-Xvlenes
Concen: 0.27 ug/l
RT: 10.37 min Scan# 1604
Delta R.T. -0.00 min
Lab File: VX002535.D
Acq: 14 Jun 2018 04:01

Instrument :
MSVOA_X
ClientSampled :

Tot Ion:106 Resp: 1527
Ion Ratio Lower Upper
106 100
91 221.7 163.4 245.2




Abundance Ion 106.00 (105.70 to 106.70): V
Ion 91.00 (90.70 to 91.70): VX0






#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 12.08 min Scan# 1886
Delta R.T. -0.00 min
Lab File: VX002535.D
Acq: 14 Jun 2018 04:01

Tgt Ion:152 Resp: 211092
Ion Ratio Lower Upper
152 100
115 55.1 40.1 120.2
150 155.2 0.0 347.2



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V





#76

1,2,3-Trichloropropane

Concen: 22.30 ug/l

RT: 11.15 min Scan# 1732

Delta R.T. -0.15 min

Lab File: VX002535.D

Acq: 14 Jun 2018 04:01

Instrument :

MSVOA_X





ClientSampleId :

Tot Ion: 75 Resp: 103298

Ion Ratio Lower Upper

75 100

77 1.2 22.8 68.3#



#78

n-propylbenzene

Concen: 0.22 ug/l

RT: 11.37 min Scan# 1768

Delta R.T. -0.00 min

Lab File: VX002535.D




Acq: 14 Jun 2018 04:01


Tgt Ion: 91 Resp: 3665

Ion Ratio Lower Upper

91 100

120 25.1 11.8 35.4









#81
 trans-1,4-Dichloro-2-butene
 Concen: 78.17 ug/l
 RT: 11.15 min Scan# 1732
 Delta R.T. 0.07 min
 Lab File: VX002535.D
 Acq: 14 Jun 2018 04:01




Instrument :
 MSVOA_X
 ClientSampled :


Tot Ion: 75 Resp: 103298
 Ion Ratio Lower Upper
 75 100
 53 0.1 86.8 130.2#
 89 0.0 38.2 57.2#



#84
 1,2,4-Trimethylbenzene
 Concen: 0.67 ug/l
 RT: 11.81 min Scan# 1841
 Delta R.T. -0.00 min
 Lab File: VX002535.D
 Acq: 14 Jun 2018 04:01

Tgt Ion: 105 Resp: 8741
 Ion Ratio Lower Upper
 105 100
 120 46.7 22.3 66.8





#95
Naphthalene
Concen: 0.24 ug/l
RT: 13.84 min Scan# 2174
Delta R.T. -0.00 min
Lab File: VX002535.D
Acq: 14 Jun 2018 04:01

Instrument :
MSVOA_X
ClientSampleId :

Tot Ion	Ratio	Lower	Upper
128	100		
127	12.7	10.4	15.6
129	11.7	8.6	13.0

