

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061318\
 Data File : VX002525.D
 Acq On : 13 Jun 2018 23:35
 Operator : JC/MD
 Sample : VX0613MBL01
 Misc : 5.00µ/10mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0613MBL01

Quant Time: Jun 14 03:49:35 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 05 03:22:35 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	203214	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	296691	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	281188	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	160489	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	162544	56.38	ug/l	0.00
Spiked Amount	50.000		Recovery	=	112.76%	
35) Dibromofluoromethane	5.51	113	121253	51.97	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.94%	
50) Toluene-d8	8.72	98	468641	52.40	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.80%	
62) 4-Bromofluorobenzene	11.14	95	164172	47.61	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.22%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061318\
Data File : VX002525.D
Acq On : 13 Jun 2018 23:35
Operator : JC/MD
Sample : VX0613MBL01
Misc : 5.00µL/10mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0613MBL01

Quant Time: Jun 14 03:49:35 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X060418W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 05 03:22:35 2018
Response via : Initial Calibration

