

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061318\
 Data File : VX002534.D
 Acq On : 14 Jun 2018 03:35
 Operator : JC/MD
 Sample : J3383-11
 Misc : 6.10g/5mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 B-40VR

Manual Integrations
 APPROVED

MMDadoda
 6/14/2018 11:29:34 AM

Quant Time: Jun 14 05:05:12 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 05 03:22:35 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	257691	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	369863	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	348618	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	213557	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	165890	44.65	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.30%	
35) Dibromofluoromethane	5.51	113	122451	42.10	ug/l	0.00
Spiked Amount	50.000		Recovery	=	84.20%	
50) Toluene-d8	8.72	98	527418	47.31	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.62%	
62) 4-Bromofluorobenzene	11.15	95	198336	46.13	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.26%	
Target Compounds						
					Qvalue	
31) Cyclohexane	5.57	56	237057	50.949	ug/l #	89
39) Methylcyclohexane	7.46	83	447389	92.906	ug/l	100
40) Benzene	6.14	78	4715	0.395	ug/l	95
67) Ethyl Benzene	10.26	91	603639	41.581	ug/l	99
68) m/p-Xylenes	10.37	106	37190	6.590	ug/l	94
73) Isopropylbenzene	11.02	105	85436	5.751	ug/l	99
78) n-propylbenzene	11.37	91	463772	28.087	ug/l	99
80) 1,3,5-Trimethylbenzene	11.51	105	497221	39.021	ug/l	100
84) 1,2,4-Trimethylbenzene	11.81	105	1698712	129.507	ug/l	98
85) sec-Butylbenzene	11.95	105	34178m	2.350	ug/l	
95) Naphthalene	13.84	128	212410	13.665	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061318\
Data File : VX002534.D
Acq On : 14 Jun 2018 03:35
Operator : JC/MD
Sample : J3383-11
Misc : 6.10g/5mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
B-40VR

Manual Integrations
APPROVED

MMDadoda
6/14/2018 11:29:34 AM

Quant Time: Jun 14 05:05:12 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X060418W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 05 03:22:35 2018
Response via : Initial Calibration

