

Data File : VX002539.D
Acq On : 14 Jun 2018 05:48
Operator : JC/MD
Sample : J3374-02ME
Misc : 6.48g/5mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EPE-109-1(80-82)ME

Quant Time: Jun 19 06:45:57 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060418W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 05 03:22:35 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	268102	50.00	ug/l	-0.01
34) 1,4-Difluorobenzene	6.86	114	383213	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	356268	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	208615	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	174731	45.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.50%	
35) Dibromofluoromethane	5.51	113	127950	42.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.92%	
50) Toluene-d8	8.72	98	546870	47.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.68%	
62) 4-Bromofluorobenzene	11.14	95	201037	45.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.26%	
Target Compounds						
68) m/p-Xylenes	10.37	106	4229	0.733	ug/l	84
84) 1,2,4-Trimethylbenzene	11.81	105	12934	1.009	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VX002539.D
Acq On : 14 Jun 2018 05:48
Operator : JC/MD
Sample : J3374-02ME
Misc : 6.48g/5mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EPE-109-1(80-82)ME

Quant Time: Jun 19 06:45:57 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060418W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 05 03:22:35 2018
Response via : Initial Calibration

