

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061418\
 Data File : VX002561.D
 Acq On : 14 Jun 2018 19:26
 Operator : JC/MD
 Sample : PB110227ZHE#05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Jun 15 01:35:22 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 05 03:22:35 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	247281	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	354204	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	331378	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	192753	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	166922	47.00	ug/l	0.00
Spiked Amount			Recovery	=	94.00%	
35) Dibromofluoromethane	5.51	113	127086	45.62	ug/l	0.00
Spiked Amount			Recovery	=	91.24%	
50) Toluene-d8	8.72	98	504876	47.29	ug/l	0.00
Spiked Amount			Recovery	=	94.58%	
62) 4-Bromofluorobenzene	11.14	95	186959	45.41	ug/l	0.00
Spiked Amount			Recovery	=	90.82%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) Chloromethane	1.32	50	636	Below Cal	#	82
5) Bromomethane	1.65	94	1222	Below Cal		95
6) Chloroethane	1.71	64	798	0.38	ug/l	70
8) Diethyl Ether	2.20	74	383	Below Cal		85
10) Methyl Iodide	2.52	142	1463	0.39	ug/l	89
11) Tert butyl alcohol	3.03	59	3546	Below Cal	#	94
13) Acrolein	2.31	56	196	Below Cal	#	14
15) Acrylonitrile	3.15	53	301	Below Cal	#	13
16) Acetone	2.45	43	26918	Below Cal		98
18) Methyl Acetate	2.78	43	23935	4.29	ug/l	99
19) Methyl tert-butyl Ether	3.21	73	135	Below Cal	#	54
20) Methylene Chloride	2.86	84	6225	Below Cal		97
22) Diisopropyl ether	3.86	45	268	Below Cal	#	35
25) 2-Butanone	4.72	43	7471	3.08	ug/l	100
28) Bromochloromethane	5.03	49	70	Below Cal	#	13
29) Tetrahydrofuran	5.17	42	1670	1.07	ug/l	78
36) 1,1-Dichloropropene	5.66	75	17540	4.71	ug/l	51
38) Carbon Tetrachloride	5.67	117	22800	5.85	ug/l	15
39) Methylcyclohexane	7.46	83	390	0.35	ug/l	87
51) 4-Methyl-2-Pentanone	8.66	43	6352	1.34	ug/l	99
65) Chlorobenzene	10.14	112	3185	0.39	ug/l	73
68) m/p-Xylenes	10.36	106	1119	0.21	ug/l	96
76) 1,2,3-Trichloropropane	11.14	75	100255	23.70	ug/l	32
81) trans-1,4-Dichloro-2-buten	11.14	75	100255	83.09	ug/l	7
84) 1,2,4-Trimethylbenzene	11.81	105	3015	0.25	ug/l	100
93) 1,2,4-Trichlorobenzene	13.65	180	1118	0.23	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061418\
Data File : VX002561.D
Acq On : 14 Jun 2018 19:26
Operator : JC/MD
Sample : PB110227ZHE#05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Jun 15 01:35:22 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X060418W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 05 03:22:35 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.00 ug/l

RT: 5.67 min Scan# 833

Delta R.T. -0.01 min

Lab File: VX002561.D

Acq: 14 Jun 2018 19:26

Instrument :

MSVOA_X


ClientSampleId :

Tot Ion:168 Resp: 247281

Ion Ratio Lower Upper




168 100

99 51.4 39.3 58.9



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



#3

Chloromethane

Concen: Below Cal

RT: 1.32 min Scan# 121

Delta R.T. 0.00 min

Lab File: VX002561.D


Acq: 14 Jun 2018 19:26

Tgt Ion: 50 Resp: 636

Ion Ratio Lower Upper



50 100

52 22.2 25.7 38.5#



Abundance Ion 49.90 (49.60 to 50.60): VX0

Ion 51.90 (51.60 to 52.60): VX0





#5

Bromomethane

Concen: Below Cal

RT: 1.65 min Scan# 174

Delta R.T. 0.01 min

Lab File: VX002561.D

Acq: 14 Jun 2018 19:26

Instrument :

MSVOA_X


ClientSampled :

Tot Ion: 94 Resp: 1222

Ion Ratio Lower Upper




94 100

96 98.7 74.9 112.3



Abundance Ion 93.90 (93.60 to 94.60): VX0

Ion 95.90 (95.60 to 96.60): VX0



#6

Chloroethane

Concen: 0.38 ug/l

RT: 1.71 min Scan# 184

Delta R.T. -0.01 min

Lab File: VX002561.D


Acq: 14 Jun 2018 19:26

Tgt Ion: 64 Resp: 798

Ion Ratio Lower Upper



64 100


66 14.9 25.5 38.3#



Abundance Ion 63.90 (63.60 to 64.60): VX0

Ion 65.90 (65.60 to 66.60): VX0









#8
Diethyl Ether
Concen: Below Cal
RT: 2.20 min Scan# 265
Delta R.T. 0.01 min
Lab File: VX002561.D
Acq: 14 Jun 2018 19:26




Instrument :
MSVOA_X
ClientSampled :


Tot Ion: 74 Resp: 383
Ion Ratio Lower Upper
74 100
45 90.9 53.1 159.4



#10
Methyl Iodide
Concen: 0.39 ug/l
RT: 2.52 min Scan# 317
Delta R.T. 0.01 min
Lab File: VX002561.D
Acq: 14 Jun 2018 19:26

Tgt Ion: 142 Resp: 1463
Ion Ratio Lower Upper
142 100
127 54.1 36.6 55.0
141 15.9 11.3 16.9









#11

Tert butyl alcohol
Concen: Below Cal
RT: 3.03 min Scan# 400
Delta R.T. -0.05 min
Lab File: VX002561.D
Acq: 14 Jun 2018 19:26

Instrument :
MSVOA_X
ClientSampled :




Tot Ion: 59 Resp: 3546
Ion Ratio Lower Upper
59 100
57 8.0 8.2 12.2#




#13

Acrolein
Concen: Below Cal
RT: 2.31 min Scan# 282
Delta R.T. 0.01 min
Lab File: VX002561.D
Acq: 14 Jun 2018 19:26

Tgt Ion: 56 Resp: 196
Ion Ratio Lower Upper
56 100
55 0.0 57.0 85.4#









#15
 Acrylonitrile
 Concen: Below Cal
 RT: 3.15 min Scan# 420
 Delta R.T. -0.01 min
 Lab File: VX002561.D
 Acq: 14 Jun 2018 19:26




Instrument :
 MSVOA_X
 ClientSampled :


Tot Ion: 53 Resp: 301
 Ion Ratio Lower Upper
 53 100
 52 195.3 66.7 100.1#
 51 40.9 29.9 44.9



#16
 Acetone
 Concen: Below Cal
 RT: 2.45 min Scan# 305
 Delta R.T. -0.01 min
 Lab File: VX002561.D
 Acq: 14 Jun 2018 19:26

Tgt Ion: 43 Resp: 26918
 Ion Ratio Lower Upper
 43 100
 58 26.8 22.1 33.1









#18
Methyl Acetate
Concen: 4.29 ug/l
RT: 2.78 min Scan# 360
Delta R.T. -0.00 min
Lab File: VX002561.D
Acq: 14 Jun 2018 19:26




Instrument :
MSVOA_X
ClientSampleId :


Tot Ion: 43 Resp: 23935
Ion Ratio Lower Upper
43 100
74 21.4 16.7 25.1



#19
Methyl tert-butyl Ether
Concen: Below Cal
RT: 3.21 min Scan# 430
Delta R.T. 0.00 min
Lab File: VX002561.D
Acq: 14 Jun 2018 19:26

Tgt Ion: 73 Resp: 135
Ion Ratio Lower Upper
73 100
57 0.0 17.7 26.5#









#20
Methvlene Chloride
Concen: Below Cal
RT: 2.86 min Scan# 373
Delta R.T. 0.00 min
Lab File: VX002561.D
Acq: 14 Jun 2018 19:26




Instrument :
MSVOA_X
ClientSampleId :


Tot Ion: 84 Resp: 6225
Ion Ratio Lower Upper
84 100
49 129.7 107.8 161.6
51 43.6 32.8 49.2
86 66.8 52.5 78.7



#22
Diisopropyl ether
Concen: Below Cal
RT: 3.86 min Scan# 537
Delta R.T. -0.02 min
Lab File: VX002561.D
Acq: 14 Jun 2018 19:26

Tgt Ion: 45 Resp: 268
Ion Ratio Lower Upper
45 100
43 0.0 46.8 70.2#
87 0.0 20.2 30.4#
59 0.0 8.7 13.1#





#25

2-Butanone

Concen: 3.08 ug/l

RT: 4.72 min Scan# 678

Delta R.T. 0.01 min

Lab File: VX002561.D

Acq: 14 Jun 2018 19:26


Instrument :
MSVOA_X
ClientSampleId :

Tot Ion: 43 Resp: 7471




Ion Ratio Lower Upper

43 100

72 23.3 18.5 27.7



Abundance Ion 43.00 (42.70 to 43.70): VX0
Ion 72.00 (71.70 to 72.70): VX0



#28

Bromochloromethane

Concen: Below Cal

RT: 5.03 min Scan# 728

Delta R.T. 0.00 min

Lab File: VX002561.D

Acq: 14 Jun 2018 19:26


Tgt Ion: 49 Resp: 70

Ion Ratio Lower Upper



49 100


129 0.0 0.0 4.2

130 0.0 61.2 91.8#



Abundance Ion 48.90 (48.60 to 49.60): VX0
Ion 128.80 (128.50 to 129.50): V
Ion 129.80 (129.50 to 130.50): V









#29
Tetrahydrofuran
Concen: 1.07 ug/l
RT: 5.17 min Scan# 752
Delta R.T. 0.01 min
Lab File: VX002561.D
Acq: 14 Jun 2018 19:26




Instrument :
MSVOA_X
ClientSampled :


Tot Ion: 42 Resp: 1670
Ion Ratio Lower Upper
42 100
72 36.8 32.0 48.0
71 13.3 30.1 45.1#



#33
1,2-Dichloroethane-d4
Concen: 47.00 ug/l
RT: 6.07 min Scan# 900
Delta R.T. 0.00 min
Lab File: VX002561.D
Acq: 14 Jun 2018 19:26

Tgt Ion: 65 Resp: 166922
Ion Ratio Lower Upper
65 100
67 51.5 0.0 102.6





#34

1,4-Difluorobenzene

Concen: 50.00 ug/l

RT: 6.87 min Scan# 1030

Delta R.T. -0.00 min

Lab File: VX002561.D

Acq: 14 Jun 2018 19:26

Instrument :

MSVOA_X

ClientSampleId :


Tot Ion:114 Resp: 354204

Ion Ratio Lower Upper

114 100

63 21.6 0.0 41.4




88 15.5 0.0 30.0



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 45.62 ug/l

RT: 5.51 min Scan# 807

Delta R.T. -0.00 min

Lab File: VX002561.D

Acq: 14 Jun 2018 19:26


Tgt Ion:113 Resp: 127086

Ion Ratio Lower Upper

113 100

111 101.8 81.4 122.0



192 23.4 19.1 28.7




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#36

1,1-Dichloropropene

Concen: 4.71 ug/l

RT: 5.66 min Scan# 832

Delta R.T. -0.15 min

Lab File: VX002561.D

Acq: 14 Jun 2018 19:26

Instrument :

MSVOA_X

ClientSampleId :


Tot Ion: 75 Resp: 17540

Ion Ratio Lower Upper

75 100

110 10.5 18.1 54.4#




77 0.4 24.3 36.5#



Abundance Ion 74.90 (74.60 to 75.60): VX002256.D (-844) (-)

Ion 109.90 (109.60 to 110.60): VX002256.D (-844) (-)

Ion 76.90 (76.60 to 77.60): VX002256.D (-844) (-)



#38

Carbon Tetrachloride

Concen: 5.85 ug/l

RT: 5.67 min Scan# 833

Delta R.T. -0.12 min

Lab File: VX002561.D

Acq: 14 Jun 2018 19:26


Tgt Ion: 117 Resp: 22800

Ion Ratio Lower Upper

117 100

119 3.8 77.4 116.0#



121 0.0 24.4 36.6#




Abundance Ion 116.80 (116.50 to 117.50): VX002256.D (-842) (-)

Ion 118.80 (118.50 to 119.50): VX002256.D (-842) (-)

Ion 120.80 (120.50 to 121.50): VX002256.D (-842) (-)









#39
Methvlcvclohexane
Concen: 0.35 ug/l
RT: 7.46 min Scan# 1127
Delta R.T. -0.01 min
Lab File: VX002561.D
Acq: 14 Jun 2018 19:26




Instrument :
MSVOA_X
ClientSampled :


Tot Ion: 83 Resp: 390
Ion Ratio Lower Upper
83 100
55 98.8 65.4 98.0#
98 49.0 37.4 56.0



#50
Toluene-d8
Concen: 47.29 ug/l
RT: 8.72 min Scan# 1334
Delta R.T. -0.00 min
Lab File: VX002561.D
Acq: 14 Jun 2018 19:26

Tgt Ion: 98 Resp: 504876
Ion Ratio Lower Upper
98 100
100 63.3 50.9 76.3





#51

4-Methyl-2-Pentanone

Concen: 1.34 ug/l

RT: 8.66 min Scan# 1324

Delta R.T. 0.01 min

Lab File: VX002561.D

Acq: 14 Jun 2018 19:26

Instrument :

MSVOA_X


ClientSampleId :

Tot Ion: 43 Resp: 6352

Ion Ratio Lower Upper




43 100

58 36.4 28.6 42.8



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 58.00 (57.70 to 58.70): VX0



#62

4-Bromofluorobenzene

Concen: 45.41 ug/l

RT: 11.14 min Scan# 1731

Delta R.T. -0.00 min

Lab File: VX002561.D

Acq: 14 Jun 2018 19:26


Tgt Ion: 95 Resp: 186959

Ion Ratio Lower Upper

95 100

174 92.8 0.0 187.4



176 91.3 0.0 181.0




Abundance Ion 94.90 (94.60 to 95.60): VX0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 10.12 min Scan# 1563
Delta R.T. -0.00 min
Lab File: VX002561.D
Acq: 14 Jun 2018 19:26

Instrument :
MSVOA_X
ClientSampleId :

Tot Ion:117 Resp: 331378
Ion Ratio Lower Upper
117 100
82 54.8 45.0 67.4
119 31.8 25.8 38.6




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VX0
Ion 118.90 (118.60 to 119.60): V






#65
Chlorobenzene
Concen: 0.39 ug/l
RT: 10.14 min Scan# 1567
Delta R.T. -0.00 min
Lab File: VX002561.D
Acq: 14 Jun 2018 19:26

Tgt Ion:112 Resp: 3185
Ion Ratio Lower Upper
112 100
114 46.7 25.3 37.9#



Abundance Ion 111.90 (111.60 to 112.60): V
Ion 113.90 (113.60 to 114.60): V









#68
m/p-Xvlenes
Concen: 0.21 ug/l
RT: 10.36 min Scan# 1603
Delta R.T. -0.01 min
Lab File: VX002561.D
Acq: 14 Jun 2018 19:26

Instrument :
MSVOA_X
ClientSampleId :

Tot Ion:106 Resp: 1119
Ion Ratio Lower Upper
106 100
91 198.7 163.4 245.2




Abundance Ion 106.00 (105.70 to 106.70): V
Ion 91.00 (90.70 to 91.70): VX0






#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 12.09 min Scan# 1886
Delta R.T. -0.00 min
Lab File: VX002561.D
Acq: 14 Jun 2018 19:26

Tgt Ion:152 Resp: 192753
Ion Ratio Lower Upper
152 100
115 56.0 40.1 120.2
150 156.4 0.0 347.2



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V





#76

1,2,3-Trichloropropane

Concen: 23.70 ug/l

RT: 11.14 min Scan# 1731

Delta R.T. -0.16 min

Lab File: VX002561.D

Acq: 14 Jun 2018 19:26

Instrument :

MSVOA_X


ClientSampleId :

Tot Ion: 75 Resp: 100255

Ion Ratio Lower Upper




75 100

77 1.0 22.8 68.3#



Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 77.00 (76.70 to 77.70): VX0



#81

trans-1,4-Dichloro-2-butene

Concen: 83.09 ug/l

RT: 11.14 min Scan# 1731

Delta R.T. 0.06 min

Lab File: VX002561.D

Acq: 14 Jun 2018 19:26


Tgt Ion: 75 Resp: 100255

Ion Ratio Lower Upper

75 100

53 0.2 86.8 130.2#



89 0.0 38.2 57.2#




Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 53.00 (52.70 to 53.70): VX0

Ion 88.90 (88.60 to 89.60): VX0









#84
 1,2,4-Trimethylbenzene
 Concen: 0.25 ug/l
 RT: 11.81 min Scan# 1841
 Delta R.T. -0.00 min
 Lab File: VX002561.D
 Acq: 14 Jun 2018 19:26

Instrument :
 MSVOA_X
 ClientSampled :

Tot Ion:105 Resp: 3015
 Ion Ratio Lower Upper
 105 100
 120 44.6 22.3 66.8



#93
 1,2,4-Trichlorobenzene
 Concen: 0.23 ug/l
 RT: 13.65 min Scan# 2142
 Delta R.T. -0.01 min
 Lab File: VX002561.D
 Acq: 14 Jun 2018 19:26

Tgt Ion:180 Resp: 1118
 Ion Ratio Lower Upper
 180 100
 182 93.4 47.7 143.1
 145 27.0 14.1 42.4

