

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061418\
 Data File : VX002568.D
 Acq On : 14 Jun 2018 23:24
 Operator : JC/MD
 Sample : VX0614WBL02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0614WBL02

Quant Time: Jun 15 03:30:59 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 05 03:22:35 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	241918	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	349280	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	323988	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	176086	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	177023	51.26	ug/l	0.00
Spiked Amount			Recovery	=	102.52%	
35) Dibromofluoromethane	5.51	113	133225	48.50	ug/l	0.00
Spiked Amount			Recovery	=	97.00%	
50) Toluene-d8	8.72	98	528609	50.21	ug/l	0.00
Spiked Amount			Recovery	=	100.42%	
62) 4-Bromofluorobenzene	11.14	95	185008	45.57	ug/l	0.00
Spiked Amount			Recovery	=	91.14%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061418\
Data File : VX002568.D
Acq On : 14 Jun 2018 23:24
Operator : JC/MD
Sample : VX0614WBL02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0614WBL02

Quant Time: Jun 15 03:30:59 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X060418W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 05 03:22:35 2018
Response via : Initial Calibration

