

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061418\
 Data File : VX002583.D
 Acq On : 15 Jun 2018 06:02
 Operator : JC/MD
 Sample : J3480-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OILY-STONES

Quant Time: Jun 15 07:39:22 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 05 03:22:35 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	223532	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	324689	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	300004	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	171130	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	155874	48.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.34%	
35) Dibromofluoromethane	5.51	113	114613	44.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.78%	
50) Toluene-d8	8.72	98	466643	47.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.36%	
62) 4-Bromofluorobenzene	11.14	95	164344	43.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.10%	
Target Compounds						
18) Methyl Acetate	2.78	43	22817	4.679	ug/l	98
20) Methylene Chloride	2.86	84	9752	1.159	ug/l	97
95) Naphthalene	13.84	128	133486	10.716	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061418\
Data File : VX002583.D
Acq On : 15 Jun 2018 06:02
Operator : JC/MD
Sample : J3480-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OILY-STONES

Quant Time: Jun 15 07:39:22 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X060418W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 05 03:22:35 2018
Response via : Initial Calibration

