

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061518\
 Data File : VX002613.D
 Acq On : 15 Jun 2018 22:49
 Operator : JC/MD
 Sample : J3516-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IDW

Quant Time: Jun 16 11:38:13 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 15 14:54:36 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	203975	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	291860	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	274245	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	155393	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	141670	45.70	ug/l	0.00
Spiked Amount			Recovery	=	91.40%	
35) Dibromofluoromethane	5.51	113	106476	45.71	ug/l	0.00
Spiked Amount			Recovery	=	91.42%	
50) Toluene-d8	8.72	98	420761	47.57	ug/l	0.00
Spiked Amount			Recovery	=	95.14%	
62) 4-Bromofluorobenzene	11.14	95	146853	43.82	ug/l	0.00
Spiked Amount			Recovery	=	87.64%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
30) Chloroform	5.21	83	14862	3.47	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061518\
Data File : VX002613.D
Acq On : 15 Jun 2018 22:49
Operator : JC/MD
Sample : J3516-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IDW

Quant Time: Jun 16 11:38:13 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 15 14:54:36 2018
Response via : Initial Calibration

