

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061518\
 Data File : VX002615.D
 Acq On : 15 Jun 2018 23:42
 Operator : JC/MD
 Sample : J3467-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 C-MW-05-061118

Quant Time: Jun 16 11:39:24 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:37:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	205509	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	295444	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	277175	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	157596	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	140079	44.85	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.70%	
35) Dibromofluoromethane	5.51	113	106797	45.29	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.58%	
50) Toluene-d8	8.72	98	423257	47.28	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.56%	
62) 4-Bromofluorobenzene	11.14	95	148646	43.81	ug/l	0.00
Spiked Amount	50.000		Recovery	=	87.62%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061518\
Data File : VX002615.D
Acq On : 15 Jun 2018 23:42
Operator : JC/MD
Sample : J3467-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
C-MW-05-061118

Quant Time: Jun 16 11:39:24 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:37:33 2018
Response via : Initial Calibration

