

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061518\
 Data File : VX002621.D
 Acq On : 16 Jun 2018 02:21
 Operator : JC/MD
 Sample : J3467-11
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PR-MW-03-061218

Quant Time: Jun 16 04:46:48 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:37:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	232738	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	334099	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	307561	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	175606	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	154622	43.72	ug/l	0.00
Spiked Amount			Recovery	=	87.44%	
35) Dibromofluoromethane	5.51	113	120964	45.37	ug/l	0.00
Spiked Amount			Recovery	=	90.74%	
50) Toluene-d8	8.72	98	467832	46.21	ug/l	0.00
Spiked Amount			Recovery	=	92.42%	
62) 4-Bromofluorobenzene	11.14	95	165552	43.15	ug/l	0.00
Spiked Amount			Recovery	=	86.30%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	6280	4.02	ug/l	99
67) Ethyl Benzene	10.26	91	5041	0.40	ug/l	95
68) m/p-Xylenes	10.37	106	5078	1.03	ug/l	94
69) o-Xylene	10.71	106	1030	0.21	ug/l	96
80) 1,3,5-Trimethylbenzene	11.51	105	3071	0.31	ug/l	98
84) 1,2,4-Trimethylbenzene	11.81	105	7688	0.76	ug/l	97
95) Naphthalene	13.84	128	5661	0.48	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061518\
Data File : VX002621.D
Acq On : 16 Jun 2018 02:21
Operator : JC/MD
Sample : J3467-11
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PR-MW-03-061218

Quant Time: Jun 16 04:46:48 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:37:33 2018
Response via : Initial Calibration

