

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061518\
 Data File : VX002622.D
 Acq On : 16 Jun 2018 02:48
 Operator : JC/MD
 Sample : J3467-12
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 C-MW-09-061218

Quant Time: Jun 16 11:49:18 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:37:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	228928	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	329989	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	303129	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	167131	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	155923	44.82	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.64%	
35) Dibromofluoromethane	5.51	113	119212	45.27	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.54%	
50) Toluene-d8	8.72	98	464821	46.48	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.96%	
62) 4-Bromofluorobenzene	11.14	95	161272	42.56	ug/l	0.00
Spiked Amount	50.000		Recovery	=	85.12%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061518\
Data File : VX002622.D
Acq On : 16 Jun 2018 02:48
Operator : JC/MD
Sample : J3467-12
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
C-MW-09-061218

Quant Time: Jun 16 11:49:18 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:37:33 2018
Response via : Initial Calibration

