

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061518\
 Data File : VX002633.D
 Acq On : 16 Jun 2018 08:31
 Operator : JC/MD
 Sample : J3477-11
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S36-EB-2018-1

Quant Time: Jun 18 07:14:33 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:37:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	221589	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	311210	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	284400	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	151772	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	148844	44.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.40%	
35) Dibromofluoromethane	5.51	113	113995	45.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.80%	
50) Toluene-d8	8.72	98	445025	47.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.38%	
62) 4-Bromofluorobenzene	11.14	95	148721	41.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.24%	
Target Compounds						
16) Acetone	2.45	43	13572	9.14	ug/l	100
20) Methylene Chloride	2.86	84	137637	45.18	ug/l	98
25) 2-Butanone	4.71	43	23924	12.26	ug/l	100
29) Tetrahydrofuran	5.17	42	1088887	851.59	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061518\
Data File : VX002633.D
Acq On : 16 Jun 2018 08:31
Operator : JC/MD
Sample : J3477-11
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S36-EB-2018-1

Quant Time: Jun 18 07:14:33 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:37:33 2018
Response via : Initial Calibration

