

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061518\
 Data File : VX002635.D
 Acq On : 16 Jun 2018 09:24
 Operator : JC/MD
 Sample : J3449-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 W-21

Quant Time: Jun 18 07:21:32 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:37:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	225736	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	321311	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	296909	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	162353	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	152603	44.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.98%	
35) Dibromofluoromethane	5.51	113	118314	46.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.28%	
50) Toluene-d8	8.72	98	457309	46.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.94%	
62) 4-Bromofluorobenzene	11.14	95	157154	42.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.18%	
Target Compounds						
16) Acetone	2.45	43	3729	2.46	ug/l #	89
30) Chloroform	5.22	83	29222	6.16	ug/l	98
95) Naphthalene	13.84	128	3774	0.35	ug/l #	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061518\
Data File : VX002635.D
Acq On : 16 Jun 2018 09:24
Operator : JC/MD
Sample : J3449-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
W-21

Quant Time: Jun 18 07:21:32 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:37:33 2018
Response via : Initial Calibration

