

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061518\
 Data File : VX002637.D
 Acq On : 16 Jun 2018 10:17
 Operator : JC/MD
 Sample : J3479-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP-0466-180611

Quant Time: Jun 18 16:11:04 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:37:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	212674	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	304405	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	296573	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	178753	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	148432	45.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.86%	
35) Dibromofluoromethane	5.51	113	112400	46.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.54%	
50) Toluene-d8	8.72	98	437127	47.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.78%	
62) 4-Bromofluorobenzene	11.14	95	160831	46.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.02%	
Target Compounds						
40) Benzene	6.15	78	36170	3.73	ug/l	Qvalue 100
65) Chlorobenzene	10.15	112	5066328	695.60	ug/l	95
87) 1,3-Dichlorobenzene	12.03	146	540246	88.45	ug/l	99
88) 1,4-Dichlorobenzene	12.10	146	1656738	266.58	ug/l	99
91) 1,2-Dichlorobenzene	12.40	146	11998369	1951.73	ug/l	87
93) 1,2,4-Trichlorobenzene	13.65	180	35414	7.94	ug/l	99
96) 1,2,3-Trichlorobenzene	14.02	180	25327	5.65	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061518\
Data File : VX002637.D
Acq On : 16 Jun 2018 10:17
Operator : JC/MD
Sample : J3479-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP-0466-180611

Quant Time: Jun 18 16:11:04 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:37:33 2018
Response via : Initial Calibration

