

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061518\
 Data File : VX002611.D
 Acq On : 15 Jun 2018 21:56
 Operator : JC/MD
 Sample : J3467-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-061218

Quant Time: Jun 16 11:37:09 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 15 14:54:36 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	205119	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	298845	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	277190	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	152971	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	142037	45.57	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	91.14%	
35) Dibromofluoromethane	5.51	113	107163	44.93	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	89.86%	
50) Toluene-d8	8.72	98	428431	47.31	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	94.62%	
62) 4-Bromofluorobenzene	11.14	95	147570	43.00	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	86.00%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061518\
Data File : VX002611.D
Acq On : 15 Jun 2018 21:56
Operator : JC/MD
Sample : J3467-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-061218

Quant Time: Jun 16 11:37:09 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 15 14:54:36 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.00 ug/l

RT: 5.67 min Scan# 834

Delta R.T. 0.00 min

Lab File: VX002611.D

Acq: 15 Jun 2018 21:56

Instrument :

MSVOA_X

ClientSampleId :


TB-061218

Tot Ion:168 Resp: 205119

Ion Ratio Lower Upper




168 100

99 53.3 39.3 58.9



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



#33

1,2-Dichloroethane-d4

Concen: 45.57 ug/l

RT: 6.07 min Scan# 900

Delta R.T. 0.00 min

Lab File: VX002611.D


Acq: 15 Jun 2018 21:56

Tgt Ion: 65 Resp: 142037

Ion Ratio Lower Upper



65 100


67 51.2 0.0 102.6



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0





#34

1,4-Difluorobenzene

Concen: 50.00 ug/l

RT: 6.87 min Scan# 1030

Delta R.T. 0.00 min

Lab File: VX002611.D

Acq: 15 Jun 2018 21:56

Instrument :

MSVOA_X

ClientSampleId :

TB-061218


Tot Ion:114 Resp: 298845

Ion Ratio Lower Upper

114 100

63 20.9 0.0 41.4




88 15.4 0.0 30.0



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 44.93 ug/l

RT: 5.51 min Scan# 807

Delta R.T. 0.00 min

Lab File: VX002611.D

Acq: 15 Jun 2018 21:56


Tgt Ion:113 Resp: 107163

Ion Ratio Lower Upper

113 100

111 100.8 81.4 122.0



192 23.4 19.1 28.7




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 47.31 ug/l

RT: 8.72 min Scan# 1334

Delta R.T. 0.00 min

Lab File: VX002611.D

Acq: 15 Jun 2018 21:56

Instrument :

MSVOA_X

ClientSampleId :


TB-061218

Tot Ion: 98 Resp: 428431

Ion Ratio Lower Upper




98 100

100 63.2 50.9 76.3



Abundance Ion 98.00 (97.70 to 98.70): VX0

Ion 100.00 (99.70 to 100.70): VX



#62

4-Bromofluorobenzene

Concen: 43.00 ug/l

RT: 11.14 min Scan# 1731

Delta R.T. 0.00 min

Lab File: VX002611.D

Acq: 15 Jun 2018 21:56


Tgt Ion: 95 Resp: 147570

Ion Ratio Lower Upper

95 100

174 95.6 0.0 187.4



176 92.8 0.0 181.0




Abundance Ion 94.90 (94.60 to 95.60): VX0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V





#63

Chlorobenzene-d5

Concen: 50.00 ug/l

RT: 10.12 min Scan# 1563

Delta R.T. 0.00 min

Lab File: VX002611.D

Acq: 15 Jun 2018 21:56

Instrument :

MSVOA_X

ClientSampled :

TB-061218


Tot Ion:117 Resp: 277190

Ion Ratio Lower Upper

117 100

82 55.2 45.0 67.4




119 32.8 25.8 38.6



Abundance Ion 116.90 (116.60 to 117.60): V

Ion 82.00 (81.70 to 82.70): VX0

Ion 118.90 (118.60 to 119.60): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.00 ug/l

RT: 12.09 min Scan# 1886

Delta R.T. 0.00 min

Lab File: VX002611.D

Acq: 15 Jun 2018 21:56


Tgt Ion:152 Resp: 152971

Ion Ratio Lower Upper

152 100

115 54.8 40.1 120.2

150 157.3 0.0 347.2



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V

